

\*\*\*\*\* End of Report \*\*\*\*\*

Christian Medical College, Vellore

CMC EQAS Co-Ordinator

Dr. Pamela Christudoss

*Pamela Christudoss*

Contact Number: 0416-2283102

Email: clinqc@cmcvellore.ac.in

Contact details:

Data in CMC EQAS reports is confidential

Homogeneity and Stability of the sample is passed.

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Sl.No	Analyte	Method / Principle	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE I	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	31	225.73	5.88	13.26	222 mg/dl	-0.28	4.76
2	UREA I	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	32	139.82	6.68	9.35	139 mg/dl	-0.09	3.30
3	CREATININE I	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	32	1.34	5.52	0.07	1.3 mg/dl	-0.54	0.03
4	T.BILIRUBIN I	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	36	5.06	10.42	0.53	3.4 mg/dl	-3.13	0.18
5	T-PROTEIN I	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	30	5.15	5.72	0.30	5.1 g/dl	-0.17	0.11
6	ALBUMIN I	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	27	2.89	5.54	0.16	3.3 g/dl	2.56	0.06
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Transasia / Erba	488	7.43	13.27	0.99	6.6 mg/dl	-0.84	0.09
8	CHOLESTEROL I	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	29	102.26	3.15	3.22	106 mg/dl	1.16	1.20
9	TRIGLYCERIDE I	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	31	104.47	10.21	10.67	112 mg/dl	0.71	3.83

Date of Report Published : 07/10/2020

Date of Result Entered : 15/09/2020

Group  
Constituent Chemistry II

Month September Year 2020

Lab Name PRANAAM HOSPITALS Lab No 5174