

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Dharshini Clinical Lab & ECG, Moondrumavadi, Madurai.	Certificate No:	AMS/LAB/2887260
		Issuing Date:	04/01/2020
		Page No:	1 of 2

Details of Device under Calibration (DUC):

DUC:	Calorimeter	Id. No:	DCL/MMVD/CAM/001
Make:	Coronation Instruments	Department:	Laboratory
Model:	M03	Calibrated On:	03/01/2020
Sl. No:	014	Calibration Due Suggested:	02/01/2021
Type:	Digital Photometer	DUC Condition on Receipt:	OK
Range & Least Count:	430,470,490,520,550,580,610,700 nm		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electrical Safety Analyser	Rigel/SafeTest60	PSG/CIRD/BMCC/3879/2019	20/01/2020
2	Thermo Hygrometer	HTC	QC/MD/20190218	10/02/2020

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

ESA Test Results:

SL.No	Parameter	ESA Reading	Unit	Normal Range	Condition
1	Current	0.10	Amps	1	Normal
2	Frequency	50.01	Hz	49 -50	Normal
3	Mains Voltage Line - Neutral	230.6	Volts	220 - 240	Normal
4	Mains Voltage Line - Earth	232.1	Volts	220 - 240	Normal
5	Mains Voltage Neutral - Earth	4.9	Volts	< 7	Normal

*Electrical Safety: Please check the attached log sheet


SL.No	Parameter	Measured	Standard	Correction	Unit
1	Lamp Voltage	12.0	12.0	+0.00	Volts
2	Detector Sensitivity	0.0	0.0	0.0	% Absorbance
3	430 (-0.00); 470 (0.00); 490 (+0.00); 520 (+0.00);				% Absorbance
4	550 (+0.00); 580 (+0.00); 610 (+0.00);700(+0.00) Auto Zero				% Absorbance

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)

B. S. S.



Approved by:
(Technical Manager)

B. S. S.