

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR

DL No. MDU/4205/20B,MDU/3971/21B

CIN : U33110TN2009PTC072140

## CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Dharshini Clinical Lab & ECG, Moondrumavadi, Madurai.	Certificate No:	AMS/LAB/2887263
		Issuing Date:	04/01/2020
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Micro Pipette	Id. No:	DCL/MMVD/MP/001
Make:	Lab	Department:	Laboratory
Model:	Variable	Calibrated On:	03/01/2020
Sl. No:	8097811	Calibration Due Suggested:	02/01/2021
Type:	Variable	DUC Condition on Receipt:	OK
Range & Least Count:	5to50µl		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	Original 2008863	16/03/2020
2	Thermo Hygrometer	HTC	QC/MD/20190218	10/02/2020

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

### Test Results:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	10	10	10	0
2	20	20	20	0
3	50	50	50	0

Allowable Tolerance: ±1%


### Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

### Calibrated By:

(Calibration Engineer)

*B. S. S.*



### Approved by:

(Technical Manager)

*B. S. S.*