

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Dharshini Clinical Lab & ECG, Moondrumavadi, Madurai.	Certificate No:	AMS/LAB/2887266
		Issuing Date:	04/01/2020
		Page No:	1 of 2

Details of Device under Calibration (DUC):

DUC:	Microscope	Id. No:	DCL/MMVD/MS/001
Make:	SH	Department:	Laboratory
Model:	Research Micro Scope	Calibrated On:	03/01/2020
Sl. No:	---	Calibration Due Suggested:	02/01/2021
Type:	Electric Monocular	DUC Condition on Receipt:	OK
Range & Least Count:	---		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electrical Safety Analyser	RigelSafeTest60	PSG/CIRD/BMCC/3879/2019	20/01/2020
2	Digital Lux Meter	HTC/ LX - 103	FSMS/CL/05/19-07/NN050-01	31/07/2020
3	Thermo Hygrometer	HTC	QC/MD/20190218	10/02/2020

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

ESA Test Results:

SL.No	Parameter	ESA Reading	Unit	Normal Range	Condition
1	Current	0.75	Amps	1	Normal
2	Frequency	50.5	Hz	49 -50	Normal
3	Mains Voltage Line - Neutral	221.7	Volts	220 - 240	Normal
4	Mains Voltage Line - Earth	221.4	Volts	220 - 240	Normal
5	Mains Voltage Neutral - Earth	4.0	Volts	< 7	Normal


*Electrical Safety: Please check the attached log sheet

SL.No.	Knob Position	DUC Reading (LUX)	Distance (Cm)
1	Low	280	5.0
2	Medium	3410	5.0
3	High	6860	5.0

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)



Approved by:
(Technical Manager)