

### CALIBRATION CERTIFICATE

DATE: 09-01-2020

**NAME OF CUSTOMER:**

**ARUL JOTHI X-RAY'S & LAB**  
 NO: 26, POOKADAI STREET,  
 RASIPURAM-637408.

BTS-350 CHEMISTRY SEMI AUTO ANALYZER PHOTOMETER BEARING.

S/NO: 801756552

CALIBRATIONS ON: 09-01-2020

**CALIBRATION VALUES:**

**LIGHT INTENSITY TEST**

LED	TINT	MAIN	REFERENCE
340	14	910496	181154
405	15	980515	695455
505	16	945952	551445
535	12	919952	451145
560	13	967841	662145
600	14	961561	512262
635	12	938452	551525
670	16	988511	581525


**PERISTALTIC PUMP ADJUSTMENT TEST**

VALUES	CURRENT	NEW
<u>SAMPLE VOLUME</u>	<u>44225</u>	<u>44225</u>
<u>POSITIONING</u>	<u>1502</u>	<u>1802</u>
<u>PUMB DELAY</u>	<u>2</u>	<u>2</u>

**NEXT CALIBRATION DUE ON:** 08-01-2021

**THIS INSTRUMENT IS CALIBRATED BY USING THE PERFORMANCE OF THE LIQUID CALIBRATOR ONLY**

**TESTED BY:** G DINESH

**APPROVED BY: Q.A.**  


**NOTE:** 1) Calibration Points & due date is given as per customer request 2) This certificate refer only to particular item submitted for certificate 3) The calibration results reported are valid at the time of and under the stated conditions of the measurements 4) Any error in the certificate should be brought to our lab within 45 days from the date of issue of certificate

Service Available @: Fully Auto Analyzer | Semi Auto Analyzer | Electrolyte Analyzer | Hematology Analyzer | Elisa Reader | Diagnostic Instruments

27, Pradhakshanam Road, A-Plot, 2<sup>nd</sup> Floor, Karur - 639 001. Cell : 96009 32255, E-mail : newtonbiomedicalservice@gmail.com

**TEST & CALIBRATION CERTIFICATE**

UNIT DESCRIPTION: MICROSCOPE

COMPANY: BLISCO

MODEL: BLS-III

NAME OF CUSTOMER  
& ADDRESS:

**ARUL JOTHI X-RAY'S & LAB**  
NO: 26, POOKADAI STREET,  
RASIPURAM-637408.

MICROSCOPE SERVICE: PREVENTIVE MAINTENANCE DONE

SERVICE: OK

CALIBRATION: OK


INSTRUMENT WORKING CONDITION: GOOD

CALIBRATION DATE: 09-01-2020

CALIBRATION VALIDITY: 08-01-2021

DATE: 09-01-2020

TEST BY: G. DINESH



**NOTE:** 1) Calibration Points & due date is given as per customer request 2) This certificate refer only to particular item submitted for certificate 3) The calibration results reported are valid at the time of and under the stated conditions of the measurements 4) Any error in the certificate should be brought to our lab within 45 days from the date of issue of certificate.