

View Lab Monthly Summary

Lab Name MICROBIOLOGICAL LABORATORY
 Lab No 11612
 Month October
 Year 2020
 Constituent Group Chemistry II

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

All Method Result

Print

Date of Result Entered : 10/10/2020

Date of Report Published :
09/11/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	30	162.46	5.55	9.01	165.1 mg/dl	0.29	3.29
2	UREA I	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	30	116.82	5.03	5.87	125.6 mg/dl	1.49	2.14
3	CREATININE I	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	31	4.93	5.11	0.25	5.41 mg/dl	1.90	0.09
4	T.BILIRUBIN I	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	39	3.47	10.23	0.36	3.76 mg/dl	0.82	0.11
5	T-PROTEIN I	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	32	4.69	7.51	0.35	4.71 g/dl	0.06	0.12
6	ALBUMIN I	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	29	2.56	9.80	0.25	2.5 g/dl	-0.24	0.09
7	URIC ACID I	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	35	3.39	9.00	0.30	3.4 mg/dl	0.03	0.10
8	CHOLESTEROL I	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	29	88.70	5.18	4.60	96 mg/dl	1.59	1.71
9	TRIGLYCERIDE I	DRY CHEMISTRY	JOHNSON & JOHNSON DRYCHEM SERIES	30	209.09	12.20	25.50	183 mg/dl	-1.02	9.31

SDI Range

Interpretation

Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

[Modify Address](#)
[Yearly Summary](#)[Modify Methods](#)
[Change Password](#)[Submit Lab Results](#)
[Home](#)[View Lab Monthly Summary](#)
[Sign-out](#)