

24th February 2020

Calibration Certificate


This is to certify that the Fully Automated Chemistry Analyser – Model ARCHITECT C4000 Instrument bearing **Se.No: C402131** has been calibrated on **24.02.2020** and the following checks and Calibrations were performed in the instrument.

- | | |
|-----------------------------------------|--------------------------------------|
| 1. Sample Pippettor Calibration | - Passed |
| 2. R1 Pippettor Calibration | - Passed |
| 3. R2 Pippettor Calibration | - Passed |
| 4. Carrier Transport Calibration | - Passed, No Step loss errors |

Next calibration is due on 23.02.2021.

Thanking you.

Yours Faithfully,
For Abbott Healthcare (P) Ltd



B.Shobanbabu
Technical Service Specialist