

VIEW LAB MONTHLY SUMMARY

Lab Name	MEDICITY MEDICAL LABORATORY	Details About Robust Analysis
Lab No	5948	Detail About Monthly Summary
Month	September	Detail about SDI
Year	2020	
Constituent Group	Chemistry II	

Click on the analyte to view Graphical Data

Date of Result Entered : 18/09/2020

Date of Report Published : 07/10/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Mindray	118	229.02	9.30	21.31	250 mg/dl	0.98	3.92
2	UREA I	UREASE UV / GLDH	Mindray	102	131.10	12.12	15.89	145.7 mg/dl	0.92	3.15
3	CREATININE I	JAFFE RATE-BLANKED & COMPENSATED/Alkaline Picrate	Mindray	84	1.38	11.75	0.16	1.4 mg/dl	0.12	0.04
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Mindray	74	5.08	12.07	0.61	5 mg/dl	-0.13	0.14
5	T-PROTEIN I	BIURET - colorimetric	Mindray	97	5.32	8.77	0.47	4.3 g/dl	-2.19	0.09
6	ALBUMIN I	BCG - colorimetric	Mindray	93	3.24	8.42	0.27	2.8 g/dl	-1.61	0.06
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Mindray	95	7.00	12.28	0.86	7.1 mg/dl	0.12	0.18
8	CHOLESTEROL I	CHOD-PAP	Mindray	109	107.92	9.86	10.64	115 mg/dl	0.67	2.04
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Mindray	99	96.47	12.92	12.47	77 mg/dl	-1.56	2.51

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

