

LAB MONTHLY SUMMARY

Lab Name MEDICITY MEDICAL LABORATORY

Lab No 5948

Month October

Year 2020

Constituent Chemistry II
Group

Date of Result Entered : 17/10/2020

Date of Report Published : 09/11/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Mindray	132	171.06	9.54	16.32	178 mg/dl	0.43	2.84
2	UREA I	UREASE UV / GLDH	Mindray	117	109.77	13.39	14.70	129 mg/dl	1.31	2.72
3	CREATININE I	JAFFE RATE-BLANKED & COMPENSATED/Alkaline Picrate	Mindray	89	4.62	13.35	0.62	4.7 mg/dl	0.13	0.13
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Mindray	84	3.65	10.21	0.37	4.1 mg/dl	1.21	0.08
5	T-PROTEIN I	BIURET - colorimetric	Mindray	100	5.00	11.76	0.59	4.2 g/dl	-1.36	0.12
6	ALBUMIN I	BCG - colorimetric	Mindray	101	3.05	8.82	0.27	3.2 g/dl	0.56	0.05
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Mindray	91	4.26	19.82	0.84	3.7 mg/dl	-0.66	0.18
8	CHOLESTEROL I	CHOD-PAP	Mindray	117	99.97	10.01	10.01	108 mg/dl	0.80	1.85
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Mindray	108	189.54	12.58	23.84	170 mg/dl	-0.82	4.59

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****