

LAB MONTHLY SUMMARY

Lab Name **MEDICITY MEDICAL LABORATORY** Lab No **5948**
 Month **May** Year **2020**
 Constituent **Chemistry II**
 Group

Date of Result Entered : 18/06/2020

Date of Report Published : 04/08/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Mindray	104	106.28	10.92	11.61	85 mg/dl	-1.83	2.28
2	UREA I	UREASE UV / GLDH	Mindray	86	28.32	13.37	3.79	21.9 mg/dl	-1.70	0.82
3	CREATININE I	JAFFE RATE-BLANKED & COMPENSATED/Alkaline Picrate	Mindray	71	4.52	12.26	0.55	4.7 mg/dl	0.32	0.13
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Mindray	65	2.71	13.19	0.36	3.2 mg/dl	1.37	0.09
5	T-PROTEIN I	BIURET - colorimetric	Mindray	77	5.21	8.59	0.45	4.6 g/dl	-1.36	0.10
6	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Mindray	87	6.29	13.42	0.84	6 mg/dl	-0.34	0.18
7	CHOLESTEROL I	CHOD-PAP	Mindray	94	108.35	8.80	9.53	120 mg/dl	1.22	1.97
8	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Mindray	80	238.99	10.44	24.94	262 mg/dl	0.92	5.58

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****