

LAB MONTHLY SUMMARY

Lab Name MEDICITY MEDICAL LABORATORY

Lab No 5948

Month June

Year 2020

Constituent Chemistry II
Group

Date of Result Entered : 18/06/2020

Date of Report Published : 08/08/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Mindray	113	80.48	9.61	7.73	61 mg/dl	-2.52	1.46
2	UREA I	UREASE UV / GLDH	Mindray	99	51.79	9.59	4.96	47.7 mg/dl	-0.82	1.00
3	CREATININE I	JAFFE RATE-BLANKED & COMPENSATED/Alkaline Picrate	Mindray	80	3.44	12.16	0.42	3.2 mg/dl	-0.57	0.09
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Mindray	77	5.35	15.08	0.81	3.6 mg/dl	-2.17	0.18
5	T-PROTEIN I	BIURET - colorimetric	Mindray	87	4.82	8.84	0.43	4.1 g/dl	-1.69	0.09
6	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Mindray	86	4.30	18.87	0.81	4.2 mg/dl	-0.12	0.17
7	CHOLESTEROL I	CHOD-PAP	Mindray	104	96.86	12.30	11.92	103 mg/dl	0.52	2.34
8	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric /End Point	Mindray	98	88.22	16.90	14.91	88 mg/dl	-0.01	3.01

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****