

**VIEW LAB MONTHLY SUMMARY**

Lab Name	MEDICITY MEDICAL LABORATORY	Details About Robust Analysis
Lab No	5948	Detail About Monthly Summary
Month	July	Detail about SDI
Year	2020	
Constituent Group	Chemistry II	

Click on the analyte to view Graphical Data

All Method Result
Print

Date of Result Entered : 16/07/2020

Date of Report Published : 14/08/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Mindray	112	380.25	11.62	44.20	418 mg/dl	0.85	8.35
2	UREA I	UREASE UV / GLDH	Mindray	98	93.04	11.08	10.31	96.4 mg/dl	0.33	2.08
3	CREATININE I	JAFFE RATE-BLANKED & COMPENSATED/Alkaline Picrate	Mindray	73	0.92	16.94	0.16	0.7 mg/dl	-1.42	0.04
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Mindray	73	4.08	14.15	0.58	4.5 mg/dl	0.73	0.14
5	T-PROTEIN I	BIURET - colorimetric	Mindray	92	5.37	7.40	0.40	5.6 g/dl	0.58	0.08
6	ALBUMIN I	BCG - colorimetric	Mindray	94	3.27	9.12	0.30	2.8 g/dl	-1.58	0.06
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Mindray	90	7.42	11.61	0.86	7.4 mg/dl	-0.02	0.18
8	CHOLESTEROL I	CHOD-PAP	Mindray	102	106.08	10.89	11.55	113 mg/dl	0.60	2.29
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Mindray	101	216.84	12.23	26.53	216 mg/dl	-0.03	5.28

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

