

Calibration Certificate

UNIT DESCRIPTION: **MODEL NO.:**

SR. NO.: **CERTIFICATE NO.:**

NAME OF CUSTOMER:

ADDRESS OF CUSTOMER:

CALIBRATION DATE: **CAL. VALIDITY:**

DETAILS OF EQUIPMENT USED FOR CALIBRATION

NOTE: Tick mark where as applicable.

Digital Lab Thermometer	<input type="checkbox"/>	LCD Tachometer	<input type="checkbox"/>	Digital Stop Watch	<input type="checkbox"/>
Auto Ranging R.P.M. Meter	<input type="checkbox"/>	Digital Strobe	<input checked="" type="checkbox"/>	Weight Set	<input type="checkbox"/>
Digital Temperature Indicator	<input type="checkbox"/>				

Sr. No.	Description	Certified By	Instrument Id. No.	Certificate No.	Calibration Date	Validity
01	Digi. Strobe	Autocal	ISR 01	18192128/Y/ISR 01	09.08.2018	08.08.2019

OBSERVATIONS:

Sr. No.	Instrument Speed Displayed (RPM)	Speed Reading on Strobe Meter (RPM)	Diff. In Actual and Strobe Meter RPM	Remarks
01	4000	4020	20	OK
02	3000	3020	20	OK
03	1000	0990	10	OK

- NOTE:**
1. Calibration Traceable to national standards.
 2. Overall deviation is within specified accuracy.
 3. Result is related only to the item calibrated.
 4. Maximum tolerance permissible is $\pm 1^\circ\text{C}$ & $\pm 5^\circ\text{C}$ whichever is applicable and $\text{RPM} \pm 2\% \text{ FSD}$ and $\pm 5\text{gm}$ to set weight.

REMARKS: ACCEPTED



CALIBRATED BY: VIJAY KADAM

APPROVED BY: Manager Q. A.