



CALIBRATION CERTIFICATE

Customer Details

Customer Name: A.J. Hospital & Research Centre

Address: National Highway 66 Kuntikana, Mangalore, Karnataka 575004

Contact Person: SachinShetty

Contact No.: +91 9900391163

Instrument Details

Instrument Type: VITEK 2 Compact 60

Serial No. VK2C18666

S/w Version: 8.01

Calibration Details

Calibration Date: 07.05.2020

Next Calibration Date: 06.05.2021

Calibration Report:

- | | |
|--|-----|
| 1. All the alignments are in the specified range | Yes |
| 2. All the voltages are in the specified range | Yes |
| 3. The Incubator & Cabinet temperatures are in the specified range | |
| a. The Incubator temperature is 35.5°C +/- 1°C | Yes |
| 4. The Vacuum Pump check done | Yes |
| a. The vacuum is 0.89PSIA +/- 0.04PSIA | Yes |
| b. Vacuum leak test is OK | Yes |
| 5. Optics Calibration done | Yes |
| a. ADC values for both Tx1 & Tx3 is in range | Yes |
| b. DAC values for both Tx1 & Tx3 is in range | Yes |

Authorised Signatory

Engineer's Name: K M Devakumar

Signature: 

Date: 07.05.2020

