



## CERTIFICATE OF CALIBRATION

ULR: CC254320000000035F

Certification No: MES/200117/C5

Calibrated for: AJ HOSPITAL&RESEARCH CENTRE, MANGALURU.

### DEVICE UNDER CALIBRATION

EQUIPMENT : MICRO PIPETTE

RANGE : 100-1000 $\mu$ l

SL.NO/ID : PW14440/ AJH/B/HRCLAB/MP/04

NO OF CHANNELS : 1

Date of Calibration : 17/01/2020

Calibrated at : Permanent Facility

MAKE : THERMO SCIENTIFIC

MODEL NO : FINNPIPETTE F3

Resolution of DUC: 1 $\mu$ l

Next calibration due : 16/01/2021

Condition of item: Good

Certified that the above instrument has been calibrated by trained technical person. The calibration Results attached with the certificate are authentic quantitative analysis report of the related instrument's which are calibrated and under valid traceability on the date of calibration.

Date of issue : 17/01/2020

Calibrated by :

  
PRABIN. K P

(Lab In Charge )

Approved Signatory

  
BINOI THOMAS

(Quality /Technical Manager)

## CALIBRATION REPORT

ULR: CC254320000000035F

Certification No: MES/200117/C5

Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	100	99.98	$\pm 0.819$	2
2		99.91		
3		99.86		
4		99.83		
5		100.00		
Repeatability =		0.033 $\mu\text{l}$		
Average volume =		99.914 $\mu\text{l}$		
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	500	499.97	$\pm 0.821$	2
2		500.02		
3		499.95		
4		499.91		
5		499.86		
Repeatability =		0.027 $\mu\text{l}$		
Average volume =		499.942 $\mu\text{l}$		
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	1000	999.99	$\pm 0.829$	2
2		999.94		
3		1000.12		
4		999.96		
5		999.90		
Repeatability =		0.038 $\mu\text{l}$		
Average volume =		999.982 $\mu\text{l}$		

The reported expanded uncertainty is the uncertainty in measurement multiplied by the coverage factor k corresponds to a normal distribution at a confidence level of 95.45%.

**Uncertainty Budget**

Source of uncertainty	Estimates	Unit	Exp. Uncertainty	k	Probability Distribution	Standard uncertainty	Sensitivity Coefficient	Uncertainty contribution	Degree of Freedom
Balance	995.51	mg	6.07E-05	2.00	Type B, Normal	3.036E-05	1.005E+00	3.050E-05	∞
Readability of balance	0.01	mg	5.00E-03	1.22	Type B, Rtglr	4.082E-03	1.005E+00	4.100E-03	∞
Temperature measurement	1.0045	µl/mg	1.40E-04	2.00	Type B, Normal	7.000E-05	9.955E+02	6.969E-02	∞
Thermal Exp. Coeff	0.0003	per °C	0.0003	1.73	Type B, Rtglr	1.732E-04	3.500E+02	6.062E-02	∞
Resolution of micropipette	1.0	µl	1.00E+00	1.22	Type B, Rtglr	8.164E-01	1.000E+00	8.164E-01	∞
Repeatability	999.98	µl	-	-	Type A	3.764E-02	1.000E+00	3.764E-02	4
Va	999.98	µl						8.224E-01	

Combined Standard uncertainty = 0.822 µl  
 Effective Degree of freedom = 911849  
 For the calculated DOF and at a confidence of 95.45 %, the coverage factor, k, determined from student's distribution table = 2  
**Expanded uncertainty  $U_E$ , in the determination of volume = ± 1.645 µl**

**Environmental Factors:**

Temperature: 26.5 °C ± 1 °C  
 Air pressure: 1003 mbar  
 Double distilled water,  
 Relative humidity: 55 %

**Note:**

- The results in this certificate are only related to DUC submitted calibration.
- Results Presented in this certificate & report shall not be reproduced except in full without the written approval of this centre.
- Next calibration due date mentioned as per customer requirement.
- All results reported are valid at the time of calibration and under stated condition of measurements
- The recalibration interval should be determined based on the user requirement.
- Calibration certificate issued for scientific or industrial purpose only.
- Calibration is carried out as per MES work procedure: MES/WP/VOLUME.





## CERTIFICATE OF CALIBRATION

ULR: CC25432000000036F

Certification No: MES/200117/C6

Calibrated for: AJ HOSPITAL&RESEARCH CENTRE,MANGALURU.

### DEVICE UNDER CALIBRATION

EQUIPMENT : MICRO PIPETTE

RANGE : 0.5-5ML

SL.NO/ID : PW13815/ AJH/B/HRCLAB/MP/08

NO OF CHANNELS : 1

Date of Calibration : 17/01/2020

Calibrated at : Permanent Facility

MAKE : THERMO SCIENTIFIC

MODEL NO : FINNPIPETTE F3

Resolution of DUC: 0.001 $\mu$ l

Next calibration due : 16/01/2021

Condition of item: Good

Certified that the above instrument has been calibrated by trained technical person. The calibration Results attached with the certificate are authentic quantitative analysis report of the related instrument's which are calibrated and under valid traceability on the date of calibration.

Date of issue : 17/01/2020

Calibrated by :

  
PRABIN. K P

(Lab In Charge )

Approved Signatory



BINOJ THOMAS

(Quality /Technical Manager)

## CALIBRATION REPORT

ULR: CC254320000000036F

Certification No: MES/200117/C6

Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	500	500.03	$\pm 0.151$	2.87000
2		500.09		
3		499.90		
4		499.95		
5		500.13		
Repeatability =		0.043 $\mu\text{l}$		
Average volume =		500.02 $\mu\text{l}$		
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	2500	2508.81	$\pm 6.533$	2.87000
2		2507.86		
3		2498.40		
4		2500.28		
5		2498.83		
Repeatability =		2.271 $\mu\text{l}$		
Average volume =		2502.838 $\mu\text{l}$		
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	5000	5004.31	$\pm 2.557$	2.87000
2		5002.13		
3		5000.04		
4		5000.18		
5		5000.09		
Repeatability =		0.838 $\mu\text{l}$		
Average volume =		5001.349 $\mu\text{l}$		

The reported expanded uncertainty is the uncertainty in measurement multiplied by the coverage factor k corresponds to a normal distribution at a confidence level of 95.45%.



Cert No. CC-2543



**Medical Engineering & Services**  
**Things Beyond Thoughts**  
(An ISO/IEC 17025:2005 NABL Accredited Calibration Laboratory)

Abhcon Crown,X1/411/21,Ambadi Lane  
Kokkalai,Thrissur-21. Ph: 0487 2420367  
Email: medengserv@gmail.com  
Visit us @www.mescalibration.com

## CERTIFICATE OF CALIBRATION

ULR: CC254320000000037F

Certification No: MES/200117/C7

Calibrated for: A.J HOSPITAL&RESERCH CENTRE,MANGALURU.

### DEVICE UNDER CALIBRATION

EQUIPMENT : MICRO PIPETTE

RANGE : 5-50 $\mu$ l

SL.NO/ID : OW14281/ AJH/B/HRCLAB/MP/11

NO OF CHANNELS : 1

Date of Calibration : 17/01/2020

Calibrated at : Permanent Facility

MAKE : THERMO SCIENTIFIC

MODEL NO : FINNPIPETTE F3

Resolution of DUC: 0.1 $\mu$ l

Next calibration due : 09/08/2021

Condition of item: Good

Certified that the above instrument has been calibrated by trained technical person. The calibration Results attached with the certificate are authentic quantitative analysis report of the related instrument's which are calibrated and under valid traceability on the date of calibration.

Date of issue : 10/08/2020

Calibrated by :

  
PRABIN. K P

(Lab In Charge )

Approved Signatory

  
BINJU THOMAS  
(Quality/Technical Manager)

## CALIBRATION REPORT

ULR: CC254320000000037F

Certification No: MES/200117/C7

Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	5*	4.99*	$\pm 0.095^*$	2*
2		4.95*		
3		4.90*		
4		4.94*		
5		4.85*		
Repeatability =		0.024* $\mu\text{l}$		
Average volume =		4.928* $\mu\text{l}$		
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	25	24.83	$\pm 0.134$	2.21000
2		25.10		
3		24.94		
4		24.99		
5		24.92		
Repeatability =		0.045 $\mu\text{l}$		
Average volume =		24.958 $\mu\text{l}$		
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	50	49.97	$\pm 0.135$	2.21000
2		49.94		
3		49.83		
4		50.11		
5		49.96		
Repeatability =		0.045 $\mu\text{l}$		
Average volume =		49.966 $\mu\text{l}$		

The reported expanded uncertainty is the uncertainty in measurement multiplied by the coverage factor k corresponds to a normal distribution at a confidence level of 95.45%.

**Uncertainty Budget**

Source of uncertainty	Estimates	Unit	Exp. Uncertainty	k	Probability Distribution	Standard uncertainty	Sensitivity Coefficient	Uncertainty contribution	Degree of Freedom	
Balance	49.75	mg	3.03E-06	2.00	Type B, Normal	1.517E-06	1.005E+00	1.524E-06	∞	
Readability of balance	0.01	mg	5.00E-03	1.22	Type B, Rtglr	4.082E-03	1.005E+00	4.100E-03	∞	
Temperature measurement	1.0045	μl/mg	1.40E-04	2.00	Type B, Normal	7.000E-05	4.975E+01	3.483E-03	∞	
Thermal Exp. Coeff	0.0003	per °C	0.0003	1.73	Type B, Rtglr	1.732E-04	1.749E+01	3.029E-03	∞	
Resolution of micropipette	0.1	μl	1.00E-01	1.22	Type B, Rtglr	8.164E-02	1.000E+00	8.164E-02	∞	
Repeatability	49.97	μl	-	-	Type A	4.486E-02	1.000E+00	4.486E-02	4	
Va	49.97	μl						9.335E-02		
Combined Standard uncertainty				= 0.093	μl					
Effective Degree of freedom				= 75						
For the calculated DOF and at a confidence of 95.45 %, the coverage factor, k, determined from student's distribution table = 2										
<b>Expanded uncertainty <math>U_E</math>, in the determination of volume</b>						= ±	<b>0.187</b>	μl		

**Environmental Factors:**

Temperature: 27.5 °C ± 1 °C  
 Air pressure: 1006 mbar  
 Double distilled water,  
 Relative humidity: 75 %

**Note:**

- The results in this certificate are only related to DUC submitted calibration.
- Results Presented in this certificate & report shall not be reproduced except in full without the written approval of this centre.
- Next calibration due date mentioned as per customer requirement.
- All results reported are valid at the time of calibration and under stated condition of measurements
- The recalibration interval should be determined based on the user requirement.
- Calibration certificate issued for scientific or industrial purpose only.
- Calibration is carried out as per MES work procedure: MES/WP/VOLUME.







Cert No. CC-2543



**Medical Engineering & Services**

**Things Beyond Thoughts**

(An ISO/IEC 17025:2005 NABL Accredited Calibration Laboratory)

Abhcon Crown, X1/411/21, Ambadi Lane  
Kokkalai, Thrissur-21. Ph: 0487 2420367  
Email: medengserv@gmail.com  
Visit us @www.mescalibration.com

## CERTIFICATE OF CALIBRATION

ULR: CC25432000000038F

Certification No: MES/200117/C8

Calibrated for: A.J HOSPITAL&RESERCH CENTRE,MANGALURU.

### DEVICE UNDER CALIBRATION

EQUIPMENT : MICRO PIPETTE

RANGE : 100-1000 $\mu$ l

SL.NO/ID : PW10143/ AJH/B/HRCLAB/MP/05

NO OF CHANNELS : 1

Date of Calibration : 17/01/2020

Calibrated at : Permanent Facility

MAKE : THERMO SCIENTIFIC

MODEL NO : FINNPIPETTE F3

Resolution of DUC: 1 $\mu$ l

Next calibration due : 16/01/2021

Condition of item: Good

Certified that the above instrument has been calibrated by trained technical person. The calibration Results attached with the certificate are authentic quantitative analysis report of the related instrument's which are calibrated and under valid traceability on the date of calibration.

Date of issue : 17/01/2020

Calibrated by :

  
PRABIN. K P

(Lab In Charge )

Approved Signatory

  
BINOJ THOMAS  
(Quality / Technical Manager)

## CALIBRATION REPORT

ULR: CC254320000000038F

Certification No: MES/200117/C8

Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	100	99.94	$\pm 0.817$	2
2		99.98		
3		99.90		
4		99.97		
5		99.89		
Repeatability =		0.018 $\mu\text{l}$		
Average volume =		99.934 $\mu\text{l}$		
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	500	499.86	$\pm 0.825$	2
2		499.93		
3		500.11		
4		499.96		
5		499.82		
Repeatability =		0.05 $\mu\text{l}$		
Average volume =		499.936 $\mu\text{l}$		
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	1000	1000.07	$\pm 0.828$	2
2		999.94		
3		999.96		
4		999.89		
5		999.95		
Repeatability =		0.03 $\mu\text{l}$		
Average volume =		999.962 $\mu\text{l}$		

The reported expanded uncertainty is the uncertainty in measurement multiplied by the coverage factor k corresponds to a normal distribution at a confidence level of 95.45%.

**Uncertainty Budget**

Source of uncertainty	Estimates	Unit	Exp. Uncertainty	k	Probability Distribution	Standard uncertainty	Sensitivity Coefficient	Uncertainty contribution	Degree of Freedom
Balance	995.59	mg	6.07E-05	2.00	Type B, Normal	3.037E-05	1.005E+00	3.050E-05	$\infty$
Readability of balance	0.01	mg	5.00E-03	1.22	Type B, Rtglr	4.082E-03	1.005E+00	4.100E-03	$\infty$
Temperature measurement	1.0045	$\mu$ l/mg	1.40E-04	2.00	Type B, Normal	7.000E-05	9.956E+02	6.969E-02	$\infty$
Thermal Exp. Coeff	0.0003	per °C	0.0003	1.73	Type B, Rtglr	1.732E-04	3.500E+02	6.062E-02	$\infty$
Resolution of micropipette	1.0	$\mu$ l	1.00E+00	1.22	Type B, Rtglr	8.164E-01	1.000E+00	8.164E-01	$\infty$
Repeatability	999.96	$\mu$ l	-	-	Type A	2.970E-02	1.000E+00	2.970E-02	4
Va	999.96	$\mu$ l						8.221E-01	

Combined Standard uncertainty = 0.822  $\mu$ l

Effective Degree of freedom = 2349351

For the calculated DOF and at a confidence of 95.45 %, the coverage factor, k, determined from student's distribution table = 2

**Expanded uncertainty  $U_E$ , in the determination of volume =  $\pm$  1.644  $\mu$ l**

**Environmental Factors:**

Temperature: 27.5 °C  $\pm$  1 °C

Air pressure: 1006 mbar

Double distilled water,

Relative humidity: 75 %

**Note:**

- The results in this certificate are only related to DUC submitted calibration.
- Results Presented in this certificate & report shall not be reproduced except in full without the written approval of this centre.
- Next calibration due date mentioned as per customer requirement.
- All results reported are valid at the time of calibration and under stated condition of measurements
- The recalibration interval should be determined based on the user requirement.
- Calibration certificate issued for scientific or industrial purpose only.
- Calibration is carried out as per MES work procedure: MES/WP/VOLUME.





Cert No. CC-2543



# Medical Engineering & Services

*Things Beyond Thoughts*

(An ISO/IEC 17025:2005 NABL Accredited Calibration Laboratory)

Abhcon Crown,X1/411/21,Ambadi Lane  
Kokkalai,Thrissur-21. Ph: 0487 2420367  
Email: medengserv@gmail.com  
Visit us @www.mescalibration.com

## CERTIFICATE OF CALIBRATION

ULR: CC254320000000039F

Certification No: MES/200117/C9

Calibrated for: A.J HOSPITAL&RESERCH CENTRE,MANGALURU.

### DEVICE UNDER CALIBRATION

EQUIPMENT : MICRO PIPETTE

RANGE : 5-50 $\mu$ l

SL.NO/ID : OW13463/ AJH/B/HRCLAB/MP/12

NO OF CHANNELS : 1

Date of Calibration : 17/01/2020

Calibrated at : Permanent Facility

MAKE : THERMO SCIENTIFIC

MODEL NO : FINNPIPETTE F3

Resolution of DUC: 0.1 $\mu$ l

Next calibration due : 16/01/2021

Condition of item: Good

Certified that the above instrument has been calibrated by trained technical person. The calibration Results attached with the certificate are authentic quantitative analysis report of the related instrument's which are calibrated and under valid traceability on the date of calibration.

Date of issue : 17/01/2020

Calibrated by :

PRABIN. K P

(Lab In Charge )

Approved Signatory



BINOJ THOMAS

(Quality/Technical Manager)

## CALIBRATION REPORT

ULR: CC25432000000039F

Certification No: MES/200117/C9

Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	5	4.96	$\pm 0.414$	2
2		4.83		
3		4.98		
4		4.86		
5		4.99		
Repeatability =		0.033 $\mu\text{l}$		
Average		4.926 $\mu\text{l}$		
volume =				
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	25	24.89	$\pm 0.413$	2
2		24.96		
3		25.08		
4		24.95		
5		24.92		
Repeatability =		0.033 $\mu\text{l}$		
Average		24.962 $\mu\text{l}$		
volume =				
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	50	50.10	$\pm 0.415$	2
2		49.98		
3		49.87		
4		49.95		
5		49.99		
Repeatability =		0.037 $\mu\text{l}$		
Average		49.982 $\mu\text{l}$		
volume =				

The reported expanded uncertainty is the uncertainty in measurement multiplied by the coverage factor k corresponds to a normal distribution at a confidence level of 95.45%.

**Uncertainty Budget**

Source of uncertainty	Estimates	Unit	Exp. Uncertainty	k	Probability Distribution	Standard uncertainty	Sensitivity Coefficient	Uncertainty contribution	Degree of Freedom
Balance	49.88	mg	3.04E-06	2.00	Type B, Normal	1.521E-06	1.005E+00	1.528E-06	∞
Readability of balance	0.01	mg	5.00E-03	1.22	Type B, Rtgir	4.082E-03	1.005E+00	4.100E-03	∞
Temperature measurement	1.0045	µl/mg	1.40E-04	2.00	Type B, Normal	7.000E-05	4.988E+01	3.492E-03	∞
Thermal Exp. Coeff	0.0003	per °C	0.0003	1.73	Type B, Rtgir	1.732E-04	1.749E+01	3.030E-03	∞
Resolution of micropipette	0.5	µl	5.00E-01	1.22	Type B, Rtgir	4.082E-01	1.000E+00	4.082E-01	∞
Repeatability	49.98	µl	-	-	Type A	3.723E-02	1.000E+00	3.723E-02	4
Va	49.98	µl						4.099E-01	

Combined Standard uncertainty = 0.410 µl  
 Effective Degree of freedom = 58758  
 For the calculated DOF and at a confidence of 95.45 %, the coverage factor, k, determined from student's distribution table = 2  
**Expanded uncertainty U<sub>E</sub>, in the determination of volume = ± 0.820 µl**

**Environmental Factors:**

Temperature: 27.5 °C ± 1 °C  
 Air pressure: 1006 mbar  
 Double distilled water,  
 Relative humidity: 75 %

**Note:**

- The results in this certificate are only related to DUC submitted calibration.
- Results Presented in this certificate & report shall not be reproduced except in full without the written approval of this centre.
- Next calibration due date mentioned as per customer requirement.
- All results reported are valid at the time of calibration and under stated condition of measurements
- The recalibration interval should be determined based on the user requirement.
- Calibration certificate issued for scientific or industrial purpose only.
- Calibration is carried out as per MES work procedure: MES/WP/VOLUME.





Cert No. CC-2543



**Medical Engineering & Services**  
**Things Beyond Thoughts**  
(An ISO/IEC 17025:2005 NABL Accredited Calibration Laboratory)

Abhcon Crown,X1/411/21,Ambadi Lane  
Kokkalai,Thrissur-21. Ph: 0487 2420367  
Email: medengserv@gmail.com  
Visit us @www.mescalibration.com

## CERTIFICATE OF CALIBRATION

ULR: CC25432000000040F

Certification No: MES/200117/C10

Calibrated for: A.J HOSPITAL&RESERCH CENTRE,MANGALURU.

### DEVICE UNDER CALIBRATION

EQUIPMENT : MICRO PIPETTE

RANGE : 0.5-5ML

SL.NO/ID : PW13799/ AJH/B/HRCLAB/MP/07

NO OF CHANNELS : 1

Date of Calibration : 17/01/2020

Calibrated at : Permanent Facility

MAKE : THERMO SCIENTIFIC

MODEL NO : FINNPIPETTE F3

Resolution of DUC: 0.001 $\mu$ l

Next calibration due : 16/01/2021

Condition of item: Good

Certified that the above instrument has been calibrated by trained technical person. The calibration Results attached with the certificate are authentic quantitative analysis report of the related instrument's which are calibrated and under valid traceability on the date of calibration.

Date of issue : 17/01/2020

Calibrated by :

PRABIN. KP

(Lab In Charge )

Approved Signatory

BINOJ THOMAS

(Quality /Technical Manager)

## CALIBRATION REPORT

ULR: CC25432000000040F

Certification No: MES/200117/C10

Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	500	499.89	$\pm 0.164$	2.87000
2		499.96		
3		500.13		
4		499.97		
5		499.85		
Repeatability =		0.048 $\mu\text{l}$		
Average volume =		499.96 $\mu\text{l}$		
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	2500	2497.76	$\pm 0.516$	2.87000
2		2497.87		
3		2498.22		
4		2497.66		
5		2497.78		
Repeatability =		0.097 $\mu\text{l}$		
Average volume =		2497.858 $\mu\text{l}$		
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	5000	4995.52	$\pm 1.01$	2.87000
2		4995.56		
3		4996.47		
4		4995.61		
5		4995.66		
Repeatability =		0.179 $\mu\text{l}$		
Average volume =		4995.764 $\mu\text{l}$		

The reported expanded uncertainty is the uncertainty in measurement multiplied by the coverage factor k corresponds to a normal distribution at a confidence level of 95.45%.



**Uncertainty Budget**

Source of uncertainty	Estimates	Unit	Exp. Uncertainty	k	Probability Distribution	Standard uncertainty	Sensitivity Coefficient	Uncertainty contribution	Degree of Freedom
Balance	4973.14	mg	3.03E-04	2.00	Type B, Normal	1.517E-04	1.005E+00	1.524E-04	$\infty$
Readability of balance	0.01	mg	5.00E-03	1.22	Type B, Rtglr	4.082E-03	1.005E+00	4.100E-03	$\infty$
Temperature measurement	1.0045	$\mu$ /mg	1.40E-04	2.00	Type B, Normal	7.000E-05	4.973E+03	3.481E-01	$\infty$
Thermal Exp. Coeff	0.0003	per °C	0.0003	1.73	Type B, Rtglr	1.732E-04	1.749E+03	3.029E-01	$\infty$
Resolution of micropipette	0.0	$\mu$ l	1.00E-03	1.22	Type B, Rtglr	8.164E-04	1.000E+00	8.164E-04	$\infty$
Repeatability	4995.76	$\mu$ l	-	-	Type A	1.789E-01	1.000E+00	1.789E-01	4
Va	4995.76	$\mu$ l						4.949E-01	

Combined Standard uncertainty = 0.495  $\mu$ l

Effective Degree of freedom = 234

For the calculated DOF and at a confidence of 95.45 %, the coverage factor, k, determined from student's distribution table = 2

**Expanded uncertainty  $U_E$ , in the determination of volume =  $\pm$  0.990  $\mu$ l**

**Environmental Factors:**

Temperature: 27.5 °C  $\pm$  1 °C

Air pressure: 1006 mbar

Double distilled water,

Relative humidity: 75 %

**Note:**

- The results in this certificate are only related to DUC submitted calibration.
- Results Presented in this certificate & report shall not be reproduced except in full without the written approval of this centre.
- Next calibration due date mentioned as per customer requirement.
- All results reported are valid at the time of calibration and under stated condition of measurements
- The recalibration interval should be determined based on the user requirement.
- Calibration certificate issued for scientific or industrial purpose only.
- Calibration is carried out as per MES work procedure: MES/WP/VOLUME.





Cert No. CC-2543



**Medical Engineering & Services**

**Things Beyond Thoughts**

(An ISO/IEC 17025:2005 NABL Accredited Calibration Laboratory)

Abhcon Crown, X1/411/21, Ambadi Lane  
Kokkalai, Thrissur-21. Ph: 0487 2420367  
Email: medengserv@gmail.com  
Visit us @www.mescalibration.com

## CERTIFICATE OF CALIBRATION

ULR: CC254320000000041F

Certification No: MES/200117/C11

Calibrated for: A.J HOSPITAL & RESEARCH CENTRE, MANGALURU.

### DEVICE UNDER CALIBRATION

EQUIPMENT : MICRO PIPETTE

RANGE : 100-1000 $\mu$ l

SL.NO/ID : PW14444/ AJH/B/HRCLAB/MP/03

NO OF CHANNELS : 1

Date of Calibration : 17/01/2020

Calibrated at : Permanent Facility

MAKE : THERMO SCIENTIFIC

MODEL NO : FINNPIPETTE F3

Resolution of DUC: 1 $\mu$ l

Next calibration due : 16/01/2021

Condition of item: Good

Certified that the above instrument has been calibrated by trained technical person. The calibration Results attached with the certificate are authentic quantitative analysis report of the related instrument's which are calibrated and under valid traceability on the date of calibration.

Date of issue : 17/01/2020

Calibrated by :

  
PRABIN K P

(Lab In Charge )

Approved Signatory

  
BINOJ THOMAS

(Quality / Technical Manager)

## CALIBRATION REPORT

ULR: CC254320000000041F

Certification No: MES/200117/C11

Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	100	99.98	$\pm 0.818$	2
2		99.93		
3		99.89		
4		99.94		
5		99.86		
Repeatability =		0.021 $\mu\text{l}$		
Average volume =		99.918 $\mu\text{l}$		
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	500	499.88	$\pm 0.821$	2
2		499.93		
3		499.96		
4		499.82		
5		499.99		
Repeatability =		0.03 $\mu\text{l}$		
Average volume =		499.916 $\mu\text{l}$		
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	1000	999.94	$\pm 0.828$	2
2		1000.07		
3		999.90		
4		999.93		
5		999.95		
Repeatability =		0.029 $\mu\text{l}$		
Average volume =		999.958 $\mu\text{l}$		

The reported expanded uncertainty is the uncertainty in measurement multiplied by the coverage factor k corresponds to a normal distribution at a confidence level of 95.45%.

**Uncertainty Budget**

Source of uncertainty	Estimates	Unit	Exp. Uncertainty	k	Probability Distribution	Standard uncertainty	Sensitivity Coefficient	Uncertainty contribution	Degree of Freedom
Balance	995.46	mg	6.07E-05	2.00	Type B, Normal	3.036E-05	1.005E+00	3.050E-05	∞
Readability of balance	0.01	mg	5.00E-03	1.22	Type B, Rtglr	4.082E-03	1.005E+00	4.100E-03	∞
Temperature measurement	1.0045	µl/mg	1.40E-04	2.00	Type B, Normal	7.000E-05	9.955E+02	6.968E-02	∞
Thermal Exp. Coeff	0.0003	per °C	0.0003	1.73	Type B, Rtglr	1.732E-04	3.500E+02	6.062E-02	∞
Resolution of micropipette	1.0	µl	1.00E+00	1.22	Type B, Rtglr	8.164E-01	1.000E+00	8.164E-01	∞
Repeatability	999.96	µl	-	-	Type A	2.935E-02	1.000E+00	2.935E-02	4
Va	999.96	µl						8.221E-01	

Combined Standard uncertainty = 0.822 µl  
 Effective Degree of freedom = 2460523  
 For the calculated DOF and at a confidence of 95.45 %, the coverage factor, k, determined from student's distribution table = 2  
**Expanded uncertainty  $U_E$ , in the determination of volume = ± 1.644 µl**

**Environmental Factors:**

Temperature: 27.5 ° C ± 1 ° C  
 Air pressure: 1006 mbar  
 Double distilled water,  
 Relative humidity: 75 %

**Note:**

- The results in this certificate are only related to DUC submitted calibration.
- Results Presented in this certificate & report shall not be reproduced except in full without the written approval of this centre.
- Next calibration due date mentioned as per customer requirement.
- All results reported are valid at the time of calibration and under stated condition of measurements
- The recalibration interval should be determined based on the user requirement.
- Calibration certificate issued for scientific or industrial purpose only.
- Calibration is carried out as per MES work procedure: MES/WP/VOLUME.





Cert No. CC-2543



**Medical Engineering & Services**  
**Things Beyond Thoughts**  
(An ISO/IEC 17025:2005 NABL Accredited Calibration Laboratory)

Abhcon Crown,X1/411/21,Ambadi Lane  
Kokkalai,Thrissur-21. Ph: 0487 2420367  
Email: medengserv@gmail.com  
Visit us @www.mescalibration.com

## CERTIFICATE OF CALIBRATION

ULR: CC25432000000042F

Certification No: MES/200117/C12

Calibrated for: A.J HOSPITAL&RESERCH CENTRE,MANGALURU.

### DEVICE UNDER CALIBRATION

EQUIPMENT : MICRO PIPETTE

RANGE : 100-1000 $\mu$ l

SL.NO/ID : PW14441/ AJH/B/HRCLAB/MP/01

NO OF CHANNELS : 1

Date of Calibration : 17/01/2020

Calibrated at : Permanent Facility

MAKE : THERMO SCIENTIFIC

MODEL NO : FINNPIPETTE F3

Resolution of DUC: 1  $\mu$ l

Next calibration due : 16/01/2021

Condition of item: Good

Certified that the above instrument has been calibrated by trained technical person. The calibration Results attached with the certificate are authentic quantitative analysis report of the related instrument's which are calibrated and under valid traceability on the date of calibration.

Date of issue : 17/01/2020

Calibrated by :

  
PRABIN. K.P

(Lab In Charge )

Approved Signatory

  
BINDI THOMAS

(Quality /Technical Manager)

## CALIBRATION REPORT

ULR: CC254320000000042F

Certification No: MES/200117/C12

Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	100	99.98	$\pm 0.819$	2
2		99.91		
3		99.86		
4		99.83		
5		100.00		
Repeatability =		0.033 $\mu\text{l}$		
Average volume =		99.914 $\mu\text{l}$		
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	500	499.97	$\pm 0.821$	2
2		500.02		
3		499.95		
4		499.91		
5		499.86		
Repeatability =		0.027 $\mu\text{l}$		
Average volume =		499.942 $\mu\text{l}$		
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	1000	999.99	$\pm 0.829$	2
2		999.94		
3		1000.12		
4		999.96		
5		999.90		
Repeatability =		0.038 $\mu\text{l}$		
Average volume =		999.982 $\mu\text{l}$		

The reported expanded uncertainty is the uncertainty in measurement multiplied by the coverage factor k corresponds to a normal distribution at a confidence level of 95.45%.

**Uncertainty Budget**

Source of uncertainty	Estimates	Unit	Exp. Uncertainty	k	Probability Distribution	Standard uncertainty	Sensitivity Coefficient	Uncertainty contribution	Degree of Freedom
Balance	995.51	mg	6.07E-05	2.00	Type B, Normal	3.036E-05	1.005E+00	3.050E-05	∞
Readability of balance	0.01	mg	5.00E-03	1.22	Type B, Rtglr	4.082E-03	1.005E+00	4.100E-03	∞
Temperature measurement	1.0045	μl/mg	1.40E-04	2.00	Type B, Normal	7.000E-05	9.955E+02	6.969E-02	∞
Thermal Exp. Coeff	0.0003	per °C	0.0003	1.73	Type B, Rtglr	1.732E-04	3.500E+02	6.062E-02	∞
Resolution of micropipette	1.0	μl	1.00E+00	1.22	Type B, Rtglr	8.164E-01	1.000E+00	8.164E-01	∞
Repeatability	999.98	μl	-	-	Type A	3.764E-02	1.000E+00	3.764E-02	4
Va	999.98	μl						8.224E-01	
Combined Standard uncertainty				= 0.822	μl				
Effective Degree of freedom				= 911849					
For the calculated DOF and at a confidence of 95.45 %, the coverage factor, k, determined from student's distribution table = 2									
Expanded uncertainty $U_B$ in the determination of volume				= ±	1.645	μl			

**Environmental Factors:**

Temperature: 26.5 °C ± 1 °C

Air pressure: 1003 mbar

Double distilled water,

Relative humidity: 55 %

**Note:**

- The results in this certificate are only related to DUC submitted calibration.
- Results Presented in this certificate & report shall not be reproduced except in full without the written approval of this centre.
- Next calibration due date mentioned as per customer requirement.
- All results reported are valid at the time of calibration and under stated condition of measurements
- The recalibration interval should be determined based on the user requirement.
- Calibration certificate issued for scientific or industrial purpose only.
- Calibration is carried out as per MES work procedure: MES/WP/VOLUME.





Cert No. CC-2543



**Medical Engineering & Services**  
**Things Beyond Thoughts**  
(An ISO/IEC 17025:2005 NABL Accredited Calibration Laboratory)

Abhcon Crown, X1/411/21, Ambadi Lane  
Kokkalai, Thrissur-21. Ph: 0487 2420367  
Email: medengserv@gmail.com  
Visit us @www.mescalibration.com

## CERTIFICATE OF CALIBRATION

ULR: CC25432000000043F

Certification No: MES/200117/C13

Calibrated for: A.J HOSPITAL&RESERCH CENTRE,MANGALURU.

### DEVICE UNDER CALIBRATION

EQUIPMENT : MICRO PIPETTE

RANGE : 100-1000 $\mu$ l

SL.NO/ID : PW17580/AJH/B/HRCLAB/MP/02

NO OF CHANNELS : 1

Date of Calibration : 17/01/2020

Calibrated at : Permanent Facility

MAKE : THERMO SCIENTIFIC

MODEL NO : FINNPIPETTE F3

Resolution of DUC: 1  $\mu$ l

Next calibration due : 16/01/2021

Condition of item: Good

Certified that the above instrument has been calibrated by trained technical person. The calibration Results attached with the certificate are authentic quantitative analysis report of the related instrument's which are calibrated and under valid traceability on the date of calibration.

Date of issue : 17/01/2020

Calibrated by :

  
PRABIN. K P

(Lab In Charge )

Approved Signatory

  
BINOJ THOMAS

(Quality /Technical Manager)



## CALIBRATION REPORT

ULR: CC25432000000043F

Certification No: MES/200117/C13

Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	100	100.00	$\pm 0.818$	2
2		99.99		
3		99.92		
4		99.86		
5		99.91		
Repeatability =		0.026 $\mu\text{l}$		
Average volume =		99.934 $\mu\text{l}$		
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	500	499.96	$\pm 0.821$	2
2		499.89		
3		500.02		
4		499.96		
5		499.87		
Repeatability =		0.027 $\mu\text{l}$		
Average volume =		499.94 $\mu\text{l}$		
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	1000	1001.05	$\pm 0.933$	2
2		999.99		
3		1000.07		
4		999.94		
5		999.90		
Repeatability =		0.218 $\mu\text{l}$		
Average volume =		1000.191 $\mu\text{l}$		

The reported expanded uncertainty is the uncertainty in measurement multiplied by the coverage factor k corresponds to a normal distribution at a confidence level of 95.45%.

Uncertainty Budget

Source of uncertainty	Estimates	Unit	Exp. Uncertainty	k	Probability Distribution	Standard uncertainty	Sensitivity Coefficient	Uncertainty contribution	Degree of Freedom
Balance	996.57	mg	6.08E-05	2.00	Type B, Normal	3.040E-05	1.005E+00	3.053E-05	∞
Readability of balance	0.01	mg	5.00E-03	1.22	Type B, Rtglr	4.082E-03	1.005E+00	4.100E-03	∞
Temperature measurement	1.0045	μl/mg	1.40E-04	2.00	Type B, Normal	7.000E-05	9.966E+02	6.976E-02	∞
Thermal Exp. Coeff	0.0003	per °C	0.0003	1.73	Type B, Rtglr	1.732E-04	3.501E+02	6.063E-02	∞
Resolution of micropipette	1	μl	1.00E+00	1.22	Type B, Rtglr	8.164E-01	1.000E+00	8.164E-01	∞
Repeatability	1000.19	μl	-	-	Type A	2.178E-01	1.000E+00	2.178E-01	4
Va	1000.19	μl						8.500E-01	

Combined Standard uncertainty = 0.850 μl

Effective Degree of freedom = 927

For the calculated DOF and at a confidence of 95.45 %, the coverage factor, k, determined

from student's distribution table = 2

Expanded uncertainty  $U_E$  in the determination of volume = ± 1.700 μl

Environmental Factors:

Temperature: 26.5 °C ± 1 °C


Air pressure: 1003 mbar

Double distilled water,

Relative humidity: 55 %

**Note:**

- The results in this certificate are only related to DUC submitted calibration.
- Results Presented in this certificate & report shall not be reproduced except in full without the written approval of this centre.
- Next calibration due date mentioned as per customer requirement.
- All results reported are valid at the time of calibration and under stated condition of measurements
- The recalibration interval should be determined based on the user requirement.
- Calibration certificate issued for scientific or industrial purpose only.





Cert No. CC-2543



**Medical Engineering & Services**

**Things Beyond Thoughts**

(An ISO/IEC 17025:2005 NABL Accredited Calibration Laboratory)

Abhcon Crown,X1/411/21,Ambadi Lane  
Kokkalai,Thrissur-21. Ph: 0487 2420367  
Email: medengserv@gmail.com  
Visit us @www.mescalibration.com

## CERTIFICATE OF CALIBRATION

ULR: CC25432000000044F

Certification No: MES/200117/C14

Calibrated for: A.J HOSPITAL&RESERCH CENTRE,MANGALURU.

### DEVICE UNDER CALIBRATION

EQUIPMENT : MICRO PIPETTE

RANGE : 0.5-5ML

SL.NO/ID : PW13792/AJH/B/HRCLAB/MP/06

NO OF CHANNELS : 1

Date of Calibration : 17/01/2020

Calibrated at : Permanent Facility

MAKE : THERMO SCIENTIFIC

MODEL NO : FINNPIPETTE F3

Resolution of DUC:0.001  $\mu$ l

Next calibration due : 16/01/2021

Condition of item: Good

Certified that the above instrument has been calibrated by trained technical person. The calibration Results attached with the certificate are authentic quantitative analysis report of the related instrument's which are calibrated and under valid traceability on the date of calibration.

Date of issue : 17/01/2020

Calibrated by :

  
PRABIN. K P

(Lab In Charge )

Approved Signatory

  
BINOJ THOMAS  
(Quality /Technical Manager)

## CALIBRATION REPORT

ULR: CC25432000000044F

Certification No: MES/200117/C14

Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	500	499.97	$\pm 0.167$	2.87000
2		500.11		
3		499.88		
4		500.15		
5		500.07		
Repeatability =		0.049 $\mu\text{l}$		
Average		500.036 $\mu\text{l}$		
volume =				
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	2500	2497.76	$\pm 2.054$	2.87000
2		2497.92		
3		2501.38		
4		2497.70		
5		2498.28		
Repeatability =		0.699 $\mu\text{l}$		
Average		2498.607 $\mu\text{l}$		
volume =				
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	5000	4995.52	$\pm 0.997$	2.87000
2		4995.66		
3		4996.47		
4		4995.65		
5		4995.73		
Repeatability =		0.17 $\mu\text{l}$		
Average		4995.806 $\mu\text{l}$		
volume =				

The reported expanded uncertainty is the uncertainty in measurement multiplied by the coverage factor k corresponds to a normal distribution at a confidence level of 95.45%.

**Uncertainty Budget**

Source of uncertainty	Estimates	Unit	Exp. Uncertainty	k	Probability Distribution	Standard uncertainty	Sensitivity Coefficient	Uncertainty contribution	Degree of Freedom
Balance	4973.14	mg	3.03E-04	2.00	Type B, Normal	1.517E-04	1.005E+00	1.524E-04	∞
Readability of balance	0.01	mg	5.00E-03	1.22	Type B, Rtglr	4.082E-03	1.005E+00	4.100E-03	∞
Temperature measurement	1.0045	µl/mg	1.40E-04	2.00	Type B, Normal	7.000E-05	4.973E+03	3.481E-01	∞
Thermal Exp. Coeff	0.0003	per °C	0.0003	1.73	Type B, Rtglr	1.732E-04	1.749E+03	3.029E-01	∞
Resolution of micropipette	0.001	µl	1.00E-03	1.22	Type B, Rtglr	8.164E-04	1.000E+00	8.164E-04	∞
Repeatability	4995.81	µl	-	-	Type A	1.702E-01	1.000E+00	1.702E-01	4
Va	4995.81	µl						4.918E-01	

Combined Standard uncertainty = 0.492 µl  
 Effective Degree of freedom = 279  
 For the calculated DOF and at a confidence of 95.45 %, the coverage factor, k, determined from student's distribution table = 2  
**Expanded uncertainty  $U_E$  in the determination of volume = ± 0.984 µl**

**Environmental Factors:**

Temperature: 26.5 ° C ± 1 ° C  
 Air pressure: 1003 mbar  
 Double distilled water,  
 Relative humidity: 55 %

**Note:**

- The results in this certificate are only related to DUC submitted calibration.
- Results Presented in this certificate & report shall not be reproduced except in full without the written approval of this centre.
- Next calibration due date mentioned as per customer requirement.
- All results reported are valid at the time of calibration and under stated condition of measurements
- The recalibration interval should be determined based on the user requirement.
- Calibration certificate issued for scientific or industrial purpose only.





Cert No. CC-2543



**Medical Engineering & Services**

**Things Beyond Thoughts**

(An ISO/IEC 17025:2005 NABL Accredited Calibration Laboratory)

Abhcon Crown,X1/411/21,Ambadi Lane  
Kokkalai,Thrissur-21. Ph: 0487 2420367  
Email: medengserv@gmail.com  
Visit us @www.mescalibration.com

## CERTIFICATE OF CALIBRATION

ULR: CC25432000000045F

Certification No: MES/200117/C15

Calibrated for: A.J HOSPITAL&RESERCH CENTRE,MANGALURU.

### DEVICE UNDER CALIBRATION

EQUIPMENT : MICRO PIPETTE

RANGE : 5-50 $\mu$ l

SL.NO/ID : PW10158/AJH/B/HRCLAB/MP/09

NO OF CHANNELS : 1

Date of Calibration : 17/01/2020

Calibrated at : Permanent Facility

MAKE : THERMO SCIENTIFIC

MODEL NO : FINNPIPETTE F3

Resolution of DUC:0.1  $\mu$ l

Next calibration due : 16/01/2021

Condition of item: Good

Certified that the above instrument has been calibrated by trained technical person. The calibration Results attached with the certificate are authentic quantitative analysis report of the related instrument's which are calibrated and under valid traceability on the date of calibration.

Date of issue : 17/01/2020

Calibrated by :

  
PRABIN. K P

(Lab In Charge )

Approved Signatory

  
BINOI THOMAS

(Quality /Technical Manager)

## CALIBRATION REPORT

ULR: CC25432000000045F

Certification No: MES/200117/C15

Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	5	5.06	$\pm 0.414$	2
2		5.14		
3		5.01		
4		5.21		
5		5.07		
Repeatability =		0.035 $\mu\text{l}$		
Average volume =		5.101 $\mu\text{l}$		
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	25	24.92	$\pm 0.413$	2
2		24.87		
3		25.02		
4		24.94		
5		25.02		
Repeatability =		0.029 $\mu\text{l}$		
Average volume =		24.956 $\mu\text{l}$		
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	50	49.98	$\pm 0.416$	2
2		50.06		
3		49.90		
4		50.03		
5		50.13		
Repeatability =		0.039 $\mu\text{l}$		
Average volume =		50.024 $\mu\text{l}$		

The reported expanded uncertainty is the uncertainty in measurement multiplied by the coverage factor k corresponds to a normal distribution at a confidence level of 95.45%.

**Uncertainty Budget**

Source of uncertainty	Estimates	Unit	Exp. Uncertainty	k	Probability Distribution	Standard uncertainty	Sensitivity Coefficient	Uncertainty contribution	Degree of Freedom
Balance	49.76	mg	3.04E-06	2.00	Type B, Normal	1.518E-06	1.005E+00	1.525E-06	∞
Readability of balance	0.01	mg	5.00E-03	1.22	Type B, Rtglr	4.082E-03	1.005E+00	4.100E-03	∞
Temperature measurement	1.0045	µl/mg	1.40E-04	2.00	Type B, Normal	7.000E-05	4.976E+01	3.483E-03	∞
Thermal Exp. Coeff	0.0003	per °C	0.0003	1.73	Type B, Rtglr	1.732E-04	1.751E+01	3.033E-03	∞
Resolution of micropipette	0.5	µl	5.00E-01	1.22	Type B, Rtglr	4.082E-01	1.000E+00	4.082E-01	∞
Repeatability	50.02	µl	-	-	Type A	3.877E-02	1.000E+00	3.877E-02	4
Va	50.02	µl						4.101E-01	

Combined Standard uncertainty = 0.410 µl  
 Effective Degree of freedom = 50035  
 For the calculated DOF and at a confidence of 95.45 %, the coverage factor, k, determined from student's distribution table = 2  
**Expanded uncertainty  $U_E$  in the determination of volume = ± 0.820 µl**

**Environmental Factors:**

Temperature: 26.5 ° C ± 1 ° C  
 Air pressure: 1003 mbar  
 Double distilled water,  
 Relative humidity: 55 %

**Note:**

- The results in this certificate are only related to DUC submitted calibration.
- Results Presented in this certificate & report shall not be reproduced except in full without the written approval of this centre.
- Next calibration due date mentioned as per customer requirement.
- All results reported are valid at the time of calibration and under stated condition of measurements
- The recalibration interval should be determined based on the user requirement.
- Calibration certificate issued for scientific or industrial purpose only.







Cert No. CC-2543



**Medical Engineering & Services**

**Things Beyond Thoughts**

(An ISO/IEC 17025:2005 NABL Accredited Calibration Laboratory)

Abhcon Crown, X1/411/21, Ambadi Lane  
Kokkalai, Thrissur-21. Ph: 0487 2420367  
Email: medengserv@gmail.com  
Visit us @www.mescalibration.com

## CERTIFICATE OF CALIBRATION

ULR: CC25432000000046F

Certification No: MES/200117/C16

Calibrated for: A.J HOSPITAL & RESERCH CENTRE, MANGALURU.

### DEVICE UNDER CALIBRATION

EQUIPMENT : MICRO PIPETTE

RANGE : 5-50 $\mu$ l

SL.NO/ID : PW10161/AJH/B/HRCLAB/MP/10

NO OF CHANNELS : 1

Date of Calibration : 17/01/2020

Calibrated at : Permanent Facility

MAKE : THERMO SCIENTIFIC

MODEL NO : FINNPIPETTE F3

Resolution of DUC: 0.1  $\mu$ l

Next calibration due : 16/01/2021

Condition of item: Good

Certified that the above instrument has been calibrated by trained technical person. The calibration Results attached with the certificate are authentic quantitative analysis report of the related instrument's which are calibrated and under valid traceability on the date of calibration.

Date of issue : 17/01/2020

Calibrated by :

PRABIN. K P

(Lab In Charge )

Approved Signatory

BINOJ THOMAS

(Quality / Technical Manager)

## CALIBRATION REPORT

ULR: CC25432000000046F

Certification No: MES/200117/C16

Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	5	4.96	$\pm 0.103$	2
2		4.98		
3		4.83		
4		5.01		
5		4.95		
Repeatability =		0.031 $\mu\text{l}$		
Average volume =		4.948 $\mu\text{l}$		
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	25	24.88	$\pm 0.113$	2.13000
2		24.96		
3		25.08		
4		24.92		
5		24.97		
Repeatability =		0.034 $\mu\text{l}$		
Average volume =		24.964 $\mu\text{l}$		
Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	50	50.11	$\pm 0.126$	2.17000
2		49.97		
3		49.93		
4		49.88		
5		50.04		
Repeatability =		0.041 $\mu\text{l}$		
Average volume =		49.99 $\mu\text{l}$		

The reported expanded uncertainty is the uncertainty in measurement multiplied by the coverage factor k corresponds to a normal distribution at a confidence level of 95.45%.

**Uncertainty Budget**

Source of uncertainty	Estimates	Unit	Exp. Uncertainty	k	Probability Distribution	Standard uncertainty	Sensitivity Coefficient	Uncertainty contribution	Degree of Freedom
Balance	49.89	mg	3.04E-06	2.00	Type B, Normal	1.522E-06	1.005E+00	1.528E-06	∞
Readability of balance	0.01	mg	5.00E-03	1.22	Type B, Rtglr	4.082E-03	1.005E+00	4.100E-03	∞
Temperature measurement	1.0045	μl/mg	1.40E-04	2.00	Type B, Normal	7.000E-05	4.989E+01	3.492E-03	∞
Thermal Exp. Coeff	0.0003	per °C	0.0003	1.73	Type B, Rtglr	1.732E-04	1.750E+01	3.030E-03	∞
Resolution of micropipette	0.1	μl	1.00E-01	1.22	Type B, Rtglr	8.164E-02	1.000E+00	8.164E-02	∞
Repeatability	49.99	μl	-	-	Type A	4.075E-02	1.000E+00	4.075E-02	4
Va	49.99	μl						9.145E-02	

Combined Standard uncertainty = 0.091 μl

Effective Degree of freedom = 101

For the calculated DOF and at a confidence of 95.45 %, the coverage factor, k, determined from student's distribution table = 2

Expanded uncertainty  $U_B$  in the determination of volume = ± 0.183 μl

**Environmental Factors:**

Temperature: 26.5 °C ± 1 °C


Air pressure: 1003 mbar

Double distilled water,

Relative humidity: 55 %

**Note:**

- The results in this certificate are only related to DUC submitted calibration.
- Results Presented in this certificate & report shall not be reproduced except in full without the written approval of this centre.
- Next calibration due date mentioned as per customer requirement.
- All results reported are valid at the time of calibration and under stated condition of measurements.
- The recalibration interval should be determined based on the user requirement.
- Calibration certificate issued for scientific or industrial purpose only.





Cert No. CC-2543



**Medical Engineering & Services**

**Things Beyond Thoughts**

(An ISO/IEC 17025:2005 NABL Accredited Calibration Laboratory)

Abhcon Crown,X1/411/21,Ambadi Lane  
Kokkalai,Thrissur-21. Ph: 0487 2420367  
Email: medengserv@gmail.com  
Visit us @www.mescalibration.com

## CERTIFICATE OF CALIBRATION

ULR: CC254320000000047F

Certification No: MES/200117/C17

Calibrated for: A.J HOSPITAL&RESERCH CENTRE,MANGALURU.

### DEVICE UNDER CALIBRATION

EQUIPMENT : MICRO PIPETTE

RANGE : 250 $\mu$ l

SL.NO/ID : 175852237/AJH/B/HRCLAB/MP/13

NO OF CHANNELS : 1

Date of Calibration : 17/01/2020

Calibrated at : Permanent Facility

MAKE : BIOMERIUX

MODEL NO : NA

Resolution of DUC:0 $\mu$ l

Next calibration due : 16/01/2021

Condition of item: Good

Certified that the above instrument has been calibrated by trained technical person. The calibration Results attached with the certificate are authentic quantitative analysis report of the related instrument's which are calibrated and under valid traceability on the date of calibration.

Date of issue : 17/01/2020

Calibrated by :

  
PRABIN. KP

(Lab In Charge )

Approved Signatory

  
BINOJ THOMAS

(Quality /Technical Manager)



# CALIBRATION REPORT

ULR: CC254320000000047F

Certification No: MES/200117/C17

Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	250	249.85	$\pm 0.238$	2.87000
2		250.28		
3		249.94		
4		249.83		
5		249.99		
Repeatability =		0.081 $\mu\text{l}$		
Average volume =		249.978 $\mu\text{l}$		

The reported expanded uncertainty is the uncertainty in measurement multiplied by the coverage factor k corresponds to a normal distribution at a confidence level of 95.45%.

**Uncertainty Budget**

Source of uncertainty	Estimates	Unit	Exp. Uncertainty	k	Probability Distribution	Standard uncertainty	Sensitivity Coefficient	Uncertainty contribution	Degree of Freedom
Balance	248.73	mg	1.52E-05	2.00	Type B, Normal	7.586E-06	1.005E+00	7.620E-06	∞
Readability of balance	0.01	mg	5.00E-03	1.22	Type B, Rtglr	4.082E-03	1.005E+00	4.100E-03	∞
Temperature measurement	1.0045	µl/mg	1.40E-04	2.00	Type B, Normal	7.000E-05	2.487E+02	1.741E-02	∞
Thermal Exp. Coeff	0.0003	per °C	0.0003	1.73	Type B, Rtglr	1.732E-04	8.749E+01	1.515E-02	∞
Resolution of micropipette	0.0	µl	0.00E+00	1.22	Type B, Rtglr	0.000E+00	1.000E+00	0.000E+00	∞
Repeatability	249.98	µl	-	-	Type A	8.132E-02	1.000E+00	8.132E-02	4
Va	249.98	µl						8.463E-02	

Combined Standard uncertainty = 0.085 µl

Effective Degree of freedom = 5

For the calculated DOF and at a confidence of 95.45 %, the coverage factor, k, determined from student's distribution table = 2.87

Expanded uncertainty  $U_E$  in the determination of volume = ± 0.243 µl

**Environmental Factors:**

Temperature: 26.5 °C ± 1 °C

Air pressure: 1003 mbar

Double distilled water,

Relative humidity: 55 %

**Note:**

- The results in this certificate are only related to DUC submitted calibration.
- Results Presented in this certificate & report shall not be reproduced except in full without the written approval of this centre.
- Next calibration due date mentioned as per customer requirement.
- All results reported are valid at the time of calibration and under stated condition of measurements
- The recalibration interval should be determined based on the user requirement.
- Calibration certificate issued for scientific or industrial purpose only.





Cert No. CC-2543



**Medical Engineering & Services**

**Things Beyond Thoughts**

(An ISO/IEC 17025:2005 NABL Accredited Calibration Laboratory)

Abhcon Crown, X1/411/21, Ambadi Lane  
Kokkalai, Thrissur-21. Ph: 0487 2420367  
Email: medengserv@gmail.com  
Visit us @www.mescalibration.com

## CERTIFICATE OF CALIBRATION

ULR: CC25432000000048F

Certification No: MES/200117/C18

Calibrated for: A.J HOSPITAL & RESEARCH CENTRE, MANGALURU.

### DEVICE UNDER CALIBRATION

EQUIPMENT : MICRO PIPETTE

RANGE : 145 $\mu$ l

SL.NO/ID : 17586245/AJH/B/HRCLAB/MP/14

NO OF CHANNELS : 1

Date of Calibration : 17/01/2020

Calibrated at : Permanent Facility

MAKE : BIOMERIUUX

MODEL NO : NA

Resolution of DUC: 0 $\mu$ l

Next calibration due : 16/01/2021

Condition of item: Good

Certified that the above instrument has been calibrated by trained technical person. The calibration Results attached with the certificate are authentic quantitative analysis report of the related instrument's which are calibrated and under valid traceability on the date of calibration.

Date of issue : 17/01/2020

Calibrated by :

PRABIN. K P

(Lab In Charge )

Approved Signatory

  
BINJOY THOMAS

(Quality / Technical Manager)

# CALIBRATION REPORT

**ULR: CC25432000000048F**

**Certification No: MES/200117/C18**

Sl.no.	Volume $\mu\text{l}$	Actual volume $\mu\text{l}$	Expanded uncertainty $\mu\text{l}$	Coverage factor (k)
1	145	144.83	$\pm 0.129$	2.87000
2		144.97		
3		145.04		
4		144.98		
5		144.82		
Repeatability =		0.044 $\mu\text{l}$		
Average volume =		144.927 $\mu\text{l}$		

The reported expanded uncertainty is the uncertainty in measurement multiplied by the coverage factor k corresponds to a normal distribution at a confidence level of 95.45%.



**Uncertainty Budget**

Source of uncertainty	Estimates	Unit	Exp. Uncertainty	k	Probability Distribution	Standard uncertainty	Sensitivity Coefficient	Uncertainty contribution	Degree of Freedom
Balance	144.18	mg	8.79E-06	2.00	Type B, Normal	4.397E-06	1.005E+00	4.417E-06	∞
Readability of balance	0.01	mg	5.00E-03	1.22	Type B, Rtglr	4.082E-03	1.005E+00	4.100E-03	∞
Temperature measurement	1.0045	µl/mg	1.40E-04	2.00	Type B, Normal	7.000E-05	1.442E+02	1.009E-02	∞
Thermal Exp. Coeff	0.0003	per °C	0.0003	1.73	Type B, Rtglr	1.732E-04	5.072E+01	8.786E-03	∞
Resolution of micropipette	0.0	µl	0.00E+00	1.22	Type B, Rtglr	0.000E+00	1.000E+00	0.000E+00	∞
Repeatability	144.93	µl	-	-	Type A	4.395E-02	1.000E+00	4.395E-02	4
Va	144.93	µl						4.612E-02	

Combined Standard uncertainty = 0.046 µl

Effective Degree of freedom = 5

For the calculated DOF and at a confidence of 95.45 %, the coverage factor, k, determined from student's distribution table = 2.87

Expanded uncertainty  $U_E$  in the determination of volume = ± 0.132 µl

**Environmental Factors:**

Temperature: 26.5 °C ± 1 °C

Air pressure: 1003 mbar

Double distilled water,

Relative humidity: 55 %

**Note:**

- The results in this certificate are only related to DUC submitted calibration.
- Results Presented in this certificate & report shall not be reproduced except in full without the written approval of this centre.
- Next calibration due date mentioned as per customer requirement.
- All results reported are valid at the time of calibration and under stated condition of measurements.
- The recalibration interval should be determined based on the user requirement.
- Calibration certificate issued for scientific or industrial purpose only.
- Calibration is carried out as per MES work procedure: MES/WP/VOLUME.

