

LAB MONTHLY SUMMARY

Lab Name **Srinivasa X-Ray & Clinical Laboratory**Lab No **11730**Month **September**Year **2020**Constituent Group **Chemistry II**Date of Result Entered : **18/09/2020**Date of Report Published : **07/10/2020**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Transasia / Erba	595	228.07	9.56	21.80	245 mg/dl	0.78	1.79
2	UREA I	UREASE UV / GLDH	Transasia / Erba	507	130.30	14.18	18.47	155 mg/dl	1.34	1.64
3	CREATININE I	ENZYMATIC COLORIMETRIC	Transasia / Erba	119	1.38	10.97	0.15	1.6 mg/dl	1.46	0.03
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Transasia / Erba	462	5.08	9.40	0.48	5.1 mg/dl	0.04	0.04
5	T-PROTEIN I	BIURET - colorimetric	Transasia / Erba	461	5.53	8.77	0.48	5.49 g/dl	-0.08	0.05
6	ALBUMIN I	BCG - colorimetric	Transasia / Erba	435	3.27	7.77	0.25	3.3 g/dl	0.12	0.02
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Transasia / Erba	488	7.43	13.27	0.99	6.6 mg/dl	-0.84	0.09
8	CHOLESTEROL I	CHOD-PAP	Transasia / Erba	496	105.27	11.87	12.49	106 mg/dl	0.06	1.12
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Transasia / Erba	469	86.44	21.15	18.28	77 mg/dl	-0.52	1.69

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Co-Ordinator
Christian Medical College, Vellore

******* End of Report *******