

LAB MONTHLY SUMMARY

Lab Name **Srinivasa X-Ray & Clinical Laboratory**Lab No **11730**Month **August**Year **2020**Constituent Group **Chemistry II**Date of Result Entered : **20/08/2020**Date of Report Published : **10/09/2020**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Transasia / Erba	562	112.38	9.73	10.94	116 mg/dl	0.33	0.92
2	UREA I	UREASE UV / GLDH	Transasia / Erba	516	74.44	12.44	9.26	77.4 mg/dl	0.32	0.82
3	CREATININE I	ENZYMATIC COLORIMETRIC	Transasia / Erba	123	9.85	9.35	0.92	10.5 mg/dl	0.71	0.17
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Transasia / Erba	436	5.76	11.22	0.65	6.9 mg/dl	1.76	0.06
5	T-PROTEIN I	BIURET - colorimetric	Transasia / Erba	420	5.58	7.80	0.44	5.4 g/dl	-0.41	0.04
6	ALBUMIN I	BCG - colorimetric	Transasia / Erba	420	3.26	8.58	0.28	3.3 g/dl	0.14	0.03
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Transasia / Erba	431	5.90	12.36	0.73	5.9 mg/dl	0.00	0.07
8	CHOLESTEROL I	CHOD-PAP	Transasia / Erba	470	106.66	11.59	12.36	104 mg/dl	-0.22	1.14
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Transasia / Erba	458	261.34	11.29	29.52	274 mg/dl	0.43	2.76

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Co-Ordinator
Christian Medical College, Vellore

******* End of Report *******