

LAB MONTHLY SUMMARY

Lab Name **Srinivasa X-Ray & Clinical Laboratory**Lab No **11730**Month **July**Year **2020**Constituent Group **Chemistry II**Date of Result Entered : **16/07/2020**Date of Report Published : **14/08/2020**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Transasia / Erba	574	390.61	11.17	43.64	427 mg/dl	0.83	3.64
2	UREA I	UREASE UV / GLDH	Transasia / Erba	518	91.03	13.34	12.14	101 mg/dl	0.82	1.07
3	CREATININE I	ENZYMATIC COLORIMETRIC	Transasia / Erba	118	0.81	19.90	0.16	0.75 mg/dl	-0.37	0.03
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Transasia / Erba	458	4.04	12.45	0.50	4.1 mg/dl	0.12	0.05
5	T-PROTEIN I	BIURET - colorimetric	Transasia / Erba	422	5.55	9.07	0.50	5.7 g/dl	0.30	0.05
6	ALBUMIN I	BCG - colorimetric	Transasia / Erba	423	3.29	9.34	0.31	3.5 g/dl	0.68	0.03
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Transasia / Erba	471	7.59	12.33	0.94	7 mg/dl	-0.63	0.09
8	CHOLESTEROL I	CHOD-PAP	Transasia / Erba	501	104.73	12.51	13.10	101 mg/dl	-0.28	1.17
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Transasia / Erba	501	203.31	14.18	28.83	210 mg/dl	0.23	2.58

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Co-Ordinator
Christian Medical College, Vellore

******* End of Report *******