

LAB MONTHLY SUMMARY

Lab Name **Srinivasa X-Ray & Clinical Laboratory**Lab No **11730**Month **June**Year **2020**Constituent Group **Chemistry II**Date of Result Entered : **25/06/2020**Date of Report Published : **08/08/2020**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Transasia / Erba	545	80.75	11.68	9.43	78 mg/dl	-0.29	0.81
2	UREA I	UREASE UV / GLDH	Transasia / Erba	500	50.22	12.64	6.35	51 mg/dl	0.12	0.57
3	CREATININE I	ENZYMATIC COLORIMETRIC	Transasia / Erba	114	3.60	9.15	0.33	3.7 mg/dl	0.30	0.06
4	T.BILIRUBIN I	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Transasia / Erba	460	5.29	15.13	0.80	5.2 mg/dl	-0.11	0.07
5	T-PROTEIN I	BIURET - colorimetric	Transasia / Erba	412	5.05	9.05	0.46	5.8 g/dl	1.64	0.05
6	ALBUMIN I	BCG - colorimetric	Transasia / Erba	404	3.02	9.46	0.29	2.8 g/dl	-0.77	0.03
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Transasia / Erba	417	4.39	16.05	0.70	3.7 mg/dl	-0.98	0.07
8	CHOLESTEROL I	CHOD-PAP	Transasia / Erba	538	95.10	16.76	15.94	91 mg/dl	-0.26	1.37
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Transasia / Erba	478	78.60	27.17	21.36	62 mg/dl	-0.78	1.95

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Co-Ordinator
Christian Medical College, Vellore

******* End of Report *******