

We hereby certify that the following instrument

Type : INCUBATOR

Serial number: 260

In NATIONAL LAB, Krishnagiri, Tamilnadu been calibrated as per SOP provided for the 'calibrating and testing the function' from the manufacturer.

Authorised Signatory



Vijayaraj .S

Service Engineer

Date of calibration : 25-11- 2019

Please note : This Certificate is valid for a period of one year from date of issue.

CALIBRATION CERTIFICATE

We hereby certify that the following instrument

Type : Hematology Analyzer Instrument

Model : Humacount 30 Ts

Serial number: 825892

STATUS PASSED CALIBRATION REPORT


DEVICE ID:
 METHOD ID: VV1000
 ENVIRONMENTAL FACTORS
 TEMP: 20°C
 AIR DENSITY: 0.001200

DESCRIPTION: VAR VOL 100-1000µl
 DATE: 05.09.2020
 REL.HUMIDITY: 50%
 CUBIC EXP: 0.0000cm³/degc.

SAMPLE			CALIBRATION STATISTICS						
VOL	NO	Wt	VOL	VOL	INACCURACY E% MEAN	IMPRECISION CV% SD	ACTUAL	TARGET	STATUS
100.0	1	100.00	100.30	100.0	100.29	0.008	1.00	0.80	passed<0.30
100.0	2	199.99	100.30	500.0	499.14	1.056	0.17	0.60	passed<0.20
100.0	3	299.99	100.30	1000.0	1000.96	1.861	0.09	0.50	passed<0.20
500.0	1	497.80	499.24						
500.0	2	996.50	500.15						
500.0	3	1493.10	498.04						
1000.0	1	1000.00	1002.90						
1000.0	2	1997.90	1000.79						
1000.0	3	2994.20	999.19						

All Pipettes are calibrated at 20°C with degassed double distilled water conforming to DIN12650

Calibrated By:



Manager (Quality Control)



Date of calibration : 10-12-2019

Please note from date of issue. : This Certificate is valid for a period of one year

CALIBRATION CERTIFICATE

We hereby certify that the following instrument

Type : Biochemistry Analyzer Instrument

Model : URA

Serial number: 601840026 IE

In NATIONAL LAB, Krishnagiri, Tamilnadu been calibrated as per SOP provided for the 'calibrating and testing the function' from the manufacturer.

Authorised Signatory


Vijayaraj S

Service Engineer

Date of calibration : 10-12-2019

Please note from date of issue. : This Certificate is valid for a period of one year

CALIBRATION REPORT

**Customer Name and Address : NATIONAL LAB,
KRISHNAGIRI.**

Instrument Details : Biochemistry Analyzer instrument

Type : URA

Serial number : 601840026 IE

Control values

Parameter	ACTUAL	TARGET	LOW	HIGH
Albumin	3.35	3.43	2.64	4.22
Creatinine	1.27	1.04	0.81	1.27
Protein	6.24	6.44	5.73	7.15
Urea	31.7	32.1	25.0	39.0
SG OT	33	35.1	27.1	43.1
SG PT	36	35.1	27.0	43.2

Calibrated by



Vijayaraj.S(B.E., Medical Electronics Engineering),

Sales and Service Engineer,

Jaiyam agency,

Next Calibration is due on December2020..

CALIBRATION CERTIFICATE

We hereby certify that the following instrument

Type : COLORIMETER

Serial number: 0606335

In National Lab, Krishnagiri, Tamilnadu been calibrated as per SOP provided for the 'calibrating and testing the function' from the manufacturer.

Authorised Signatory


Vijayaraj

Service Engineer

Date of calibration : 12- 12 - 2019

Please note : This Certificate is valid for a period of one year from date of issue.

CALIBRATION REPORT

**Customer Name and Address : NATIONAL LAB
KRISHNAGIRI.**

Control values

Parameter	ACTUAL	TARGET	LOW	HIGH
Triglycerides	165	170	150	190
Glucose	88	92	72	120
Cholesterol	215	220	200	240

Calibrated by



Vijayaraj.S (B.E., Medical Electronics Engineering),

Sales and Service Engineer,

Jaiyam agency.

Next Calibration is due on December2020..

CALIBRATION CERTIFICATE

We hereby certify that the following instrument

Type : CLINICAL CENTRIFUGE

Model : REMI

Serial number: GHLC-11027

In NATIONAL LAB, Krishnagiri, Tamilnadu been calibrated as per SOP provided for the 'calibrating and testing the function' from the manufacturer.

Authorised Signatory


Vijayaraj S

Service Engineer

Date of calibration : 10 - 12 - 2019

Please note : This Certificate is valid for a period of one year from date of issue.

CALIBRATION REPORT

**Customer Name and Address : NATIONAL LAB,
KRISHHNAGIRI.**

**Instrument Details : REMI
Type : CLINICAL CENTRIFUGE
Serial number : GHLC-11027**