

Certificate of Calibration

Customer Name: Chinmayi Hospital, Kundapurra

Model: Automated Hematology Analyzer H560


Serial no.: K110419360S8

Calibration Done Date: 12-03-2020

Next Calibration Date on or Before: 11-03-2021

Lab Incharge: Mr Pradeep

Calibration at Site performed by



H560 CAL 1		Batch No: D162534		EXP-22-04-2021		Cal Done on: 12-03-2020		Cal Due: 11-3-2020	
Parameter	WBC	RBC	HGB	MCV	PLT	MCH	PCT	HCT	
	6.23	4.77	14.5	90.2	274	30.4	0.253	46	
	6.31	4.76	14.5	89.7	271	30.4	0.244	42.7	
	6.31	4.76	14.3	89.7	259	30.1	0.234	42	
	6.37	4.75	14.3	89.8	263	30.1	0.243	42.4	
	6.36	4.77	14.5	89.8	263	30.5	0.239	42.2	
	6.28	4.79	14.5	89.7	265	30.5	0.243	42.2	
	6.48	4.81	14.5	89.9	265	30.3	0.244	42	
	6.36	4.77	14.4	89.7	261	30.4	0.242	4	
	6.31	4.76	14.4	89.8	263	30.2	0.245	42	
	6.35	4.76	14.5	89.7	264	30.4	0.243	42	
Mean	6.33	4.77	14.44	89.8	264.8	30.33	0.24	42.2	
SD	0.07	0.02	0.08	0.16	4.49	0.15	0	0.18	
CV%	1.05	0.37	0.58	0.17	1.7	0.49	2.06	0	

H560 CAL 2		Batch No: D162535		EXP-22-04-2021		Cal Done on: 12-03-2020		Cal Due: 11-3-2020	
Parameter	WBC	RBC	HGB	MCV	PLT	MCH	PCT	HCT	
	8.88	3.87	11.2	86.3	281	29	0.251	33.4	
	8.98	3.84	11.1	85.6	284	29	0.256	32.9	
	9.1	3.84	11	86	290	28.5	0.258	33.1	
	9.12	3.86	11.2	85.9	284	29	0.255	33.2	
	8.89	3.87	11.1	85.8	285	28.8	0.261	33.2	
	8.75	3.86	11.2	85.9	279	29	0.253	33.1	
	9.09	3.85	11.1	85.8	282	28.9	0.249	33	
	9.01	3.84	11.2	85.8	281	29.9	0.251	33.1	
	8.79	3.83	11	85.7	284	29	0.253	33.3	
	8.93	3.85	11.2	85.4	286	29	0.254	33	
Mean	8.95	3.85	11.13	85.8	283.8	29	0.254	33.1	
SD	0.13	0.01	0.08	0.24	3.1	0.35	0	0.15	
CV%	1.44	0.36	0.74	0.28	1.09	0.121	1.41	0.45	

*Pradeep*

LABORATORY  
 CHINMAYI HOSPITAL  
 KUNDAPURRA

*Pradeep*