



ADARSHA CALIBRATION SERVICES

NABL ACCREDITED CALIBRATION LABORATORY

Plot No. 18,19 Survey No - 118, Byraveshwara Nagar, Manganahalli Cross, Ullal Main Road,
Ullal (P), Bangalore - 560110. Ph: +91-80-2324 0578, Mob: +91-9945086171/172/173

E: acscalib@gmail.com / enquiry@adarshacalibrations.com W: www.adarshacalibrations.com



CERTIFICATE OF CALIBRATION

Page 1 of 1

Certificate Number:	ACS/20-21/S01844-09	ULR Number:	ULR-CC-27222000004261F	
Customer Name & Address:	M/s.A.J. Hospital & Research Centre, Kuntikana, NH-66, Mangaluru-575004. Karnataka-India.			
Date Of SRF:	22-09-2020	Calibrated At:	Site	
Date of Calibration:	22-09-2020	Calibration WI No:	WI/MC/10	
Next Cal Due:	21-09-2021	Date of Issue:	24-09-2020	
Environmental Conditions:	Temperature: 23.9°C	Humidity:	55% R.H	
DETAILS OF DEVICE UNDER CALIBRATION				
Nomenclature :	RPM Meter	Serial. No:	ZFCN-10595	
Make & Model :	Remi & Neya-4	Code/ID No :	AJH/B/HRCLAB/CENT/02	
Range :	50 to 6000 RPM	Resolution:	10 RPM	
Location:	HRC Lab-(Benchtop Centrifuge)			
DETAILS OF STANDARD INSTRUMENT USED			Reference Standard: SANAS TR 45-02/01	
Nomenclature:	Digital Tachometer	Serial. No:	AE.75584	
Make / Model :	Lutron & DT-1236L	Cal Due Date:	13-03-2021	
Range :	10 to 99999 RPM	Certificate No:	ACS/MC/20-21/I02284-01	
CALIBRATION RESULTS:				
Parameter: SPEED		Discipline : Mechanical		
Sl.No.	Standard Reading (RPM)	DUC Reading (RPM)	Error Obtained ± (RPM)	Expanded Measurement Uncertainty ±(RPM)
PHOTO				
1	1003	1000	-3	3.77
2	2007	2000	-7	3.77
3	3011	3000	-11	3.77
4	4015	4000	-15	3.77
5	6019	6000	-19	3.77
REMARKS & CONCLUSION:				
1.The Measurement Uncertainty is estimated at a confidence level of 95.45 % with a coverage factor k=2.0.				
2.Next Calibration Due Date mentioned as per customer requirement.				
3.Calibration Work Order Received Through "M/s, Medical Engineering & Services".				
Calibrated By		Authorized Signatory		
Deepak Kamaraj Calibration Engineer		Manjunath I Technical Manager		

