



ADARSHA CALIBRATION SERVICES

NABL ACCREDITED CALIBRATION LABORATORY

Plot No. 18,19 Survey No - 118, Byraveshwara Nagar, Manganahalli Cross, Ullal Main Road, Ullal (P), Bangalore - 560110. Ph: +91-80-2324 0578, Mob: +91-9945086171/172/173

E: acscalib@gmail.com / enquiry@adarshacalibrations.com W: www.adarshacalibrations.com

CERTIFICATE OF CALIBRATION

Page 1 of 1

Certificate Number:

ACS/20-21/S01844-08

ULR Number: ULR-CC-272220000004260F

Customer Name &

M/s.A.J. Hospital & Research Centre,

Address:

Kuntikana, NH-66, Mangaluru-575004.

Karnataka-India.

Date Of SRF:

22-09-2020

Calibrated At:

Site

Date of Calibration:

22-09-2020

Calibration WI No:

WI/MC/10

Next Cal Due:

21-09-2021

Date of Issue:

24-09-2020

Temperature:

24.4°C Humidity:

53% R.H

Environmental Conditions:

DETAILS OF DEVICE UNDER CALIBRATION

DETAILS OF STANDARD INSTRUMENT USED

Nomenclature:

RPM Meter

Serial. No:

ZFCN-09790

Make & Model:

Remi & Neya-4

Code/ID No:

AJH/B/HRCLAB/CENT/01

Range:

50 to 6000 RPM

Resolution:

10 RPM

Location:

Range:

HRC Lab-(Benchtop Centrifuge)

Reference Standard:

SANAS TR 45-02/01

Nomenclature:

Digital Tachometer

Serial. No:

AE.75584

Make / Model:

Lutron & DT-1236L 10 to 99999 RPM

Cal Due Date: Certificate No: 13-03-2021

ACS/MC/20-21/102284-01

CALIBRATION RESULTS:

Parameter: SPEED

Discipline: Mechanical

Sl.No.	Standard Reading (RPM)	DUC Reading (RPM)	Error Obtained ± (RPM)	Expanded Measuremer Uncertainty ±(RPM)
		РНОТО		
1	1004	1000	-4	3.77
2	2006	2000	-6	3.77
3	3009	3000	-9	3.77
4	4012	4000	-12	3.77
5	6015	6000	-15	3.77

REMARKS & CONCLUSION:

1. The Measurement Uncertainty is estimated at a confidence level of 95.45 % with a coverage factor k=2.0.

2. Next Calibration Due Date mentioned as per customer requirement.

3. Calibration Work Order Received Through "M/s, Medical Engineering & Services".

Calibrated By

Calibration Engineer

Authorized Signatory

1 auni

Technical Manager