

LAB MONTHLY SUMMARY

Lab Name LIFELINE LABORATORY Lab No 4119
 Month October Year 2020
 Constituent Group Chemistry II

Date of Result Entered : 06/10/2020

Date of Report Published : 09/11/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Any Analyser	910	168.35	10.59	17.82	170 mg/dl	0.09	1.18
2	UREA I	UREASE UV / GLDH	Any Analyser	550	108.37	14.86	16.10	106 mg/dl	-0.15	1.37
3	CREATININE I	JAFFE-END POINT	Any Analyser	358	4.52	13.11	0.59	4.7 mg/dl	0.30	0.06
4	T,BILIRUBIN I	EVELYN-MALLOY	Any Analyser	74	3.54	15.84	0.56	4.6 mg/dl	1.89	0.13
5	T-PROTEIN I	BIURET - colorimetric	Any Analyser	667	5.06	10.46	0.53	3.8 g/dl	-2.38	0.04
6	ALBUMIN I	BCG - colorimetric	Any Analyser	622	3.02	9.51	0.29	2.9 g/dl	-0.42	0.02
7	URIC ACID I	ENZYMATIC / URICASE Colorimetric	Any Analyser	517	4.49	18.56	0.83	3.5 mg/dl	-1.19	0.07
8	CHOLESTEROL I	CHOD-PAP	Any Analyser	771	101.13	13.52	13.68	101 mg/dl	-0.01	0.98
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser	735	180.95	13.56	24.53	174 mg/dl	-0.28	1.81

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****