



Prism Calibration Centre

www.prismcalibration.com / www.prismtesthouse.com

E-mail : prismtesthouse@gmail.com, info@prismcalibration.com

NABL Accredited Calibration Lab



CALIBRATION CERTIFICATE

Client Name:- Life Line Laboratory
First Floor, Near Santram Mandir, Nadiad, Gujarat
India 387001.

Date of Issue :- 18/03/2020
Date of Calibration :- 18/03/2020
Suggested Due Date:- 17/03/2021

FORMAT NO.:-PRISM/M&V/GW/01
Calibration Certificate No.:-200317-12-001

ISSUE NO.:-02
Date Of Receipt / JRF. No. :-17/03/2020 (200317-12)

Details of Unit Under Calibration (UUC)

Identification No.	OB490710
Name Of Instrument	Micro Pipette
Make	P Fact
Temperature	24.2°C
Rel. Humidity	48.9% RH
Baro. Pressure	1003.1 hPa

Serial No.	Volume (µl)	Weight Readings (mg)	Corrected Volume Readings@27°C (µl)	Deviation in Readings (µl)	Expanded Uncertainty (±µl)
1	200.0	194.75	195.63	-4.37	0.40
2	400.0	394.62	396.46	-3.54	0.40
3	600.0	594.37	597.07	-2.93	0.65
4	800.0	793.75	797.35	-2.65	0.65
5	1000.0	992.98	997.51	-2.49	0.65

Remarks :

(1)The Measurement Confidence Level is 95.45 % with Coverage factor k = 2.
(2)These Results Are Obtained At The Time Of Calibration Only Relate to the item Calibrated.
(3)The Calibration Certificate Shall not be reproduced except in full without written approval of Prism Calibration Centre.
(4) Each Instrument is Individually Calibrated On Electronic Balance.
(5) Calibration Procedure No. PRISM/CAL/SOP/VG/04 As Per ISO 4787 & ISO/TR 20461
(6) 750 mmHg=99.99 kPa
(7) Weight in mg or g
(8) Environment Condition During Calibration: 15 to 30°C, 35% to 85% RH.
(9)The Instrument is Calibrated Using as Standard electronic Balance With Calibration Traceability to NABL.


Details of Instrument Used for Calibration	Name	Digital Weighing Balance
	Range	0-220 g
	I.D No.	PRISM/ENV1/WBAC/509
	Serial No.	592433
	Due Date	28-06-2020

Prism Calibration Center

PCC/MV/N/1906134/01

CALIBRATED BY:-
NAME:-Ms.Suvidha Awaik *Suvidha Awaik*
DESIGNATION:-Technical Staff Member

APPROVED BY:-
NAME:-Mrs.Krishna Kinariwala *Krishna Kinariwala*
DESIGNATION:-Quality Manager



ULR NO : (CC248020000015898F)

-----End Of Certificate-----

Inst. Supply & Calibration, Validation, GPCB Approved Env. Auditor

F/101, Rudraksh Complex-II, Beside Syndicate Bank, Nr. Jasoda Nagar Cross Road-III, GIDC, Vatva, A'bad-382 445.
(M) 99255 94828, 94286 02063, 98245 26444 Customer Care : 90990 62852, 90990 62851