



**CERTIFICATE OF CALIBRATION**


ADL/MO/SER/1050/20-21

Customer Name : Nidan Pathology Laboratory  
 Address of the customer : Bhatkal Gate, Aurangabad,  
 Instrument Model : BC 3000 Plus – 3 Part Hematology Analyzer  
 Serial No. : RJ-8B127371  
 Date of Installation : 19/02/2019  
 Date of Calibration : 04/11/2020

This is to certify that the above mentioned product has been calibrated for CBC parameters (WBC, RBC, HGB, MCV, PLT) according to the standard operating procedure given by the MINDRAY

The reference instruments used for value assignment are managed by the traceability system in Shenzhen Mindray Bio – Medical Electronics Co., Ltd, and these are traceable to the International Standards such as ICSS/WHO, NCCLS.

For AGAPPE DIAGNOSTICS LTD.



Authorized Signatory *Rajan Patil*  
 Name: Rajan Patil  
 Designation: Engineer - Customer Service

**AGAPPE DIAGNOSTICS LTD.**

ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC15413

**CORPORATE OFFICE / REGENT PLANT**  
 Agappe Mills, Patholathani SPUS, SDC, Erudathuram, Kerala - 681 562, India.  
 Tel: +91 484 250 7050 | Email: agappe@agappe.in

**EQUIPMENT PLANT**  
 2/505 CB, Block No. 32, KEMFISA Small Industrial Park,  
 NRIEAI, Cochin, Kerala, India - 685 721. Tel: +91 484 276 7477.

**KOLKATA OFFICE**  
 406, Merin Matrix, Plot No-10, Block-DW, Sector V,  
 Salt Lake City, Kolkata - 700 091.  
 Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

**MUMBAI (REGISTERED OFFICE)**  
 601 & 602, 4<sup>th</sup> Floor, Sakinaka Parkside Centre, 119

**DELHI OFFICE**  
 DSM 540, 6<sup>th</sup> Floor, DLF Tower, Sakinaka

**BANGALORE OFFICE**  
 5-6, 8<sup>th</sup> Floor, Red Cross Bldg, No.26 Bann Course Road,

DETAILS OF CONTROL RUN

Instrument Model : BC 3000 PLUS – 3 Part Hematology Analyzer  
 Serial No : RJ-8B127371  
 Date of Control Run : 04/11/2020  
 Control Used : Agappe  
 Lot No : 30090077  
 Expiry : 05/12/2020

The results of repeatability test are summarized in the following table. Refer the attached control sheets for details.

Sample No	WBC K/ $\mu$ L	RBC M/ $\mu$ L	HGB g/dL	MCV fL	PLT K/ $\mu$ L
1	6.9	4.65	12.4	79.0	220
2	7.0	4.65	12.2	79.0	220
3	7.2	4.67	12.3	79.1	218
4	7.3	4.63	12.5	79.0	219
5	7.1	4.65	12.6	69.9	218
Instrument Mean	7.1	4.65	12.4	79.0	219
Range	$\pm 0.9$	$\pm 0.4$	$\pm 0.6$	$\pm 6$	$\pm 36$
CV%	0.9%	0.1%	0.3	1.2%	3%

This is to certify that the above mentioned product has been calibrated for CBC parameters (WBC, RBC, HGB, MCV, PLT) according to the standard operating procedure given by the Mindray

The control run after calibration shows good repeatability.

Next Calibration due on **03/MAY/2021**

For AGAPPE DIAGNOSTICS LTD.



Authorized Signatory

Name: Rajan Patil

Designation: Engineer - Customer Service

**AGAPPE DIAGNOSTICS LTD.**

ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC15413

CORPORATE OFFICE / REAGENT PLANT  
 AGAPPE PVT. LTD. Kalyaneshwar (P.O.) SNo. 1, Trambhavan, Kalyan - 401 502, India.  
 Tel. : +91 484 266 7676 | Email: agappe@agappe.in

EQUIPMENT PLANT  
 2/508-C2, Block No. 32, WIMERA Small Industrial Park,  
 Noida, Gurgaon, Gurgaon, India - 122 021. Tel. : +91 484 276 7477.

DELHI OFFICE  
 DDA 501, 5<sup>th</sup> Floor, DLF Tower, Shivaji Park.

KOLKATA OFFICE  
 404, Merlin Matrix, Plot No-40, Block-DW, Sector V,  
 Salt Lake City, Kolkata - 700 091.  
 Tel. : +91 33 4003 0451 | Email: kolkatoffice@agappe.in

BANGALORE OFFICE  
 5-6, B<sup>th</sup> Floor, And Cross Bhawan, No.26 Race Course Road.

MUMBAI REGISTERED OFFICE  
 419 & 420, 6<sup>th</sup> Floor, Jambhulkar Building, Chhatrapati Shivaji Maharaj, Mumbai - 400 004.