



UNIQUE CALIBRATION SOLUTIONS LLP

An ISO/IEC-17025:2017 / ACCREDITED BY NABL - GOVT OF INDIA

Tel: 0484-4858563, 0484 - 7966984, 9971328563, 9061728563

E-mail: sales@uniquecalibration.com, website: www.uniquecalibration.com



NABL
C.No.: CC-3102

UNIQUE
CALIBRATION & REPAIRS

CALIBRATION CERTIFICATE

UNIQUE LAB REORT NO(ULR NO)		CC310220000001042F	
SRF No.	088	Certificate No	UCSL/11-20/088_06

User Name : M/s MORRIS MATHIAS HOSPITAL
355, KP ROAD, Dr MATHIAS NAGER ,NAAGERCOIL,
TAMIL NADU-629001

Gauge For Calibration	CENTRIFUGE(RPM)
Make / Model No.	REMI
Range	0 to 3500 rpm
Least Count	--
Instrument Id. No.	MMH/BCF-1/014
Instrument Sr. No.	--
Location	PARAPSYCHOLOGY
Visual Inspection	OK
Material Receipt Date	--
Calibration Performed at	ONSITE
Date of Calibration	09.11.2020
Suggested Due Date of Calibration	09.11.2021
Calibration Certificate Issue Date	14.11.2020
ENVIRONMENTAL CONDITION	
Temperature	(25 ± 2)°C
Relative Humidity	(50 ± 20)%RH

Equipment & Master Used For calibration	
01. DIGITAL TACHO METER (NON-CONTACT)	
Make	LUTRON
Model No.	DT -2234 C
Certificate No.	PPL/M/1352/1
Next Due Date	23.11.2020
Traceability	Pascal
Accreditation No	CC-2899

Condition Of Equipment :	Un-Loaded
--------------------------	-----------

OBSERVATION RESULTS

SL. NO.	Nominal Value on UUC (rpm)	Observed Mean Value on MASTER (rpm)	% Error of Reading (±)	± Expanded Uncertainty At 95% Confidence Level (k=2)
1.	700	700.6	0.09	0.46 % rdg
2.	1400	1400.6	0.04	0.46 % rdg
3.	2100	2101.2	0.06	0.46 % rdg
4.	2800	2801.4	0.05	0.46 % rdg
5.	3500	3501.6	0.05	0.46 % rdg

Calibration Procedure Based On	: UCSL/SOP/01-RPM
--------------------------------	-------------------

NOTE :-

- The calibration results reported in this certificate are valid at the time of and the stated condition of measurement.
- This report should not be reproduced except in full without our prior permission in writing.
- Calibration certificate without signature are not valid.
- All our certificates are Traceable to National standard..
- UUC : Unit under calibration , GUT : Gauge under testing .

Calibrated By

Issued / Approved By
(MANU MATHEW)
Technical Manager



UNIQUE CALIBRATION SOLUTIONS LLP

An ISO/IEC-17025:2017 / ACCREDITED BY NABL - GOVT OF INDIA

Tel: 0484-4858563, 0484 - 7966984, 9971328563, 9061728563

E-mail: sales@uniquecalibration.com, website: www.uniquecalibration.com

UNIQUE
CALIBRATION SOLUTIONS



NABL
C.No.: CC-3102

CALIBRATION CERTIFICATE

UNIQUE LAB REORT NO(ULR NO) :		CC310220000001043F	
SRF No. :	088	Certificate No :	UCSL/11-20/088_07
User Name :	M/s MORRIS MATHIAS HOSPITAL 355,KP ROAD,Dr MATHIAS NAGER ,NAAGERCOIL, TAMIL NADU-629001		

Gauge For Calibration	CENTRIFUGE(RPM)	Equipment & Master Used For calibration 01. DIGITAL TACHO METER (NON-CONTACT) Make : LUTRON Model No. : DT-2234 C Certificate No. : PPL/M/1352/1 Next Due Date : 23.11.2020 Traceability : Pascal Accreditation No : CC-2899
Make / Model No.	ELECTRA	
Range	100 to 4000 rpm	
Least Count	1rpm	
Instrument Id. No.	MMH/LAB CEN-2/022	
Instrument Sr. No.	2015	
Location	BIOCHEMISTRY	
Visual Inspection	OK	
Material Receipt Date	--	
Calibration Performed at	ONSITE	
Date of Calibration	09.11.2020	
Suggested Due Date of Calibration	09.11.2021	
Calibration Certificate Issue Date	14.11.2020	
ENVIRONMENTAL CONDITION		
Temperature	(25 ± 2)°C	
Relative Humidity	(50 ± 20)%RH	

Condition Of Equipment :	Un-Loaded
--------------------------	-----------

OBSERVATION RESULTS

SL. NO.	Nominal Value on UUC (rpm)	Observed Mean Value on MASTER (rpm)	% Error of Reading (±)	± Expanded Uncertainty At 95% Confidence Level (k=2)
1.	100	100.1	0.12	1.35 % rdg
2.	500	500.4	0.08	0.46 % rdg
3.	1000	1000.8	0.08	0.46 % rdg
4.	2500	2502.6	0.10	0.46 % rdg
5.	4000	4003.8	0.10	0.46 % rdg

Calibration Procedure Based On	: UCSL/SOP/01-RPM
--------------------------------	-------------------

NOTE :-

- The calibration results reported in this certificate are valid at the time of and the stated condition of measurement.
- This report should not be reproduced except in full without our prior permission in writing.
- Calibration certificate without signature are not valid.
- All our certificates are Traceable to National standard.
- UUC : Unit under calibration , GUT : Gauge under testing .

Calibrated By

Issued / Approved By
 (MANU MATHEW)
 Technical Manager



UNIQUE CALIBRATION SOLUTIONS LLP

An ISO/IEC-17025:2017 / ACCREDITED BY NABL - GOVT OF INDIA

Tel: 0484-4858563, 0484 - 7966984, 9971328563, 9061728563

E-mail: sales@uniquecalibration.com, website: www.uniquecalibration.com



NABL
C.No.: CC-3102

UNIQUE
CALIBRATION SOLUTIONS

CALIBRATION CERTIFICATE

UNIQUE LAB REORT NO(ULR NO)	:	CC310220000001044F
SRF No.	:	088
Certificate No	:	UCSL/11-20/088_08
User Name	:	M/s MORRIS MATHIAS HOSPITAL 355.KP ROAD,Dr MATHIAS NAGER ,NAAGERCOIL, TAMIL NADU-629001

Gauge For Calibration	CENTRIFUGE(RPM)
Make / Model No.	EURO
Range	100 to 4000 rpm
Least Count	1rpm
Instrument Id. No.	MMH/LAB CEN-3/005
Instrument Sr. No.	201908113-12
Location	BIOCHEMISTRY
Visual Inspection	OK
Material Receipt Date	--
Calibration Performed at	ONSITE
Date of Calibration	09.11.2020
Suggested Due Date of Calibration	09.11.2021
Calibration Certificate Issue Date	14.11.2020
ENVIRONMENTAL CONDITION	
Temperature	(25 ± 2) ^o C
Relative Humidity	(50 ± 20)%RH

Equipment & Master Used For calibration	
01. DIGITAL TACHO METER (NON-CONTACT)	
Make	LUTRON
Model No.	DT -2234 C
Certificate No.	PPL/M/1352/1
Next Due Date	23.11.2020
Traceability	Pascal
Accreditation No	CC-2899

Condition Of Equipment :	Un-Loaded
--------------------------	-----------

OBSERVATION RESULTS

SL. NO.	Nominal Value on UUC (rpm)	Observed Mean Value on MASTER (rpm)	% Error of Reading (±)	± Expanded Uncertainty At 95% Confidence Level (k=2)
1.	100	100.0	0.04	1.35 % rdg
2.	500	500.6	0.13	0.46 % rdg
3.	1000	1001.2	0.12	0.46 % rdg
4.	2500	2502.8	0.11	0.46 % rdg
5.	4000	4004.4	0.11	0.46 % rdg

Calibration Procedure Based On	: UCSL/SOP/01-RPM
--------------------------------	-------------------

NOTE :-

1. The calibration results reported in this certificate are valid at the time of and the stated condition of measurement.
2. This report should not be reproduced except in full without our prior permission in writing.
3. Calibration certificate without signature are not valid.
4. All our certificates are Traceable to National standard..
5. UUC : Unit under calibration , GUT : Gauge under testing .

Calibrated By

Issued / Approved By
(MANU MATHEW)
Technical Manager