

**TO: SRI BABA DIAGNOSTIC CENTRE - NO: 6, Vellai chetty street,
Uthiramerur - 603 406**

CALIBRATION CERTIFICATE

Date : 13-11-2020


This is to certify that we have Calibrated the
BIO CHEMISTRY ANALYZER

Model :PMI - 5000 SA SERIAL NO: SA2F008

Calibration provided by the RANDOX

1. CALIBRATOR USED : CAL - ROCHE
LOT NO: 32413201
EXP.DATE : 2021-01-31
2. CALIBRATION DONE BY : V.RANJITH KUMAR
SERVICE ENGINEER
MICRO ENTERPRISES,
CHENNAI.
3. CALIBRATION DONE ON : 13-11-2020
4. NEXT CALIBRATION DUE ON : 12-11-2021

For MICRO ENTERPRISES,



Service Engineer

**TO: SRI BABA DIAGNOSTIC CENTRE, No, 6 VELLAI CHETTY STREET,
UTHIRAMERUR-63 406**

1. NAME OF THE EQUIPMENT: BIO CHEMISTRY ANALYSER

2. SERIAL NUMBER : SA2F008

3. MODEL : PMI-5000 SA

4. DETAILS OF CALIBRATING AGENCY : MICRO ENTERPRISES

5. QC USED: ROCHE C.F.A.S., EXP : 31-JAN-2021

6. CRITICAL PARAMETERS TO BE CALIBRATED :


FILTER	GAIN	AD VALUES	OFFSET
340	3.00	30269	100
405	3.00	32125	102
450	3.00	33456	103
510	3.00	30242	101
546	3.00	32548	102
578	3.00	33554	103
620	3.00	32546	102

Note: AD VALUES SHOULD BE FROM 20,000 to 40,000

6. CALIBRATION DONE BY : V.RANJITH KUMAR

Date : 13-11-2020

SERVICE ENGINEER



TO : SRI BABA DIAGNOSTIC CENTRE- NO:6, Vellai Chetty Street ,
Uthiramerur - 603 406

CONTROL DATA RESULT

Control used: Roche

Control Brand: Roche

Lot. No: - 324132 Expiry: 2021-01-31

BIO CHEMISTRY ANALYZER - MODEL: PMI-5000SA

BIO CHEMISTRY SERIAL NO: SA2F008

CONTROL RESULTS							
Parameter	GLU	UREA	CREATININE	TRIGLYCERIDES	CHOLESTROL	T.BILL	D.BILL
Control Acceptable Range	101.0	40.8	1.19	121	90.1	1.05	0.86
Acceptable Range	86-116	34.8 -46.8	0.98-1.40	103-139	76.6-103.6	0.87-1.23	0.65-1.06
Value 1	99	40	1.1	125	86	0.9	0.8
Value 2	100	41	1.15	126	85	0.95	0.85
Value 3	99	41	1	121	86	0.99	0.84
MEAN	96	40.67	1.1	38.8	85.7	0.946667	0.83
SD	1.15	1.15	0.15	5.29	1.15	0.09	0.05
CV %	1.20	2.84	14.10	13.64	1.35	9.53	6.38
CV % SPECIFIED	2.50	2.00	2.00	2.00	2.50	5.00	5.00
RESULT	PASSED	PASSED	PASSED	PASSED	PASSED	PASSED	PASSED

9. CALIBRATION DONE BY

: V. RANJITH KUMAR - SERVICE ENGINEER

Date : 13-11-2020

FOR MICRO ENTERPRISES

