

Dimension RXL MAX, bearing serial number: **2004080375** installed at **M/s. St.THERESA'S HOSPITAL, Hyderabad.** have been duly calibrated on **15/05/2020.**

Final calibration results are as follows:

PHOTOMETER DARK CALIBRATION:

Reference	:	9562.00 Hz
Sample (outer on)	:	9549.60 Hz
Sample (outer off)	:	9563.02 Hz

LAMP CALIBRATION:

Time on High	:	0.5 min
Low Calib level	:	49.0%
High Calib level	:	62.0%

PHOTOMETER—mAU OFFSET VALUES:

FILTER: 293nm ----	299
FILTER: 340nm ----	298
FILTER: 383nm ----	290
FILTER: 405nm ----	299
FILTER: 452nm ----	295
FILTER: 510nm ----	302
FILTER: 540nm ----	301
FILTER: 577nm ----	275
FILTER: 600nm ----	298
FILTER: 700nm ----	300

PHOTOMETER LAMP VOLTAGE:

Set at 24.00V calibrated by DVM

