



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 1348

Distribution No.: 151-D

Month/Year: August/2020

Instrument ID: MEDONIC CELL COUNTER M-16GP

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 27-11-2020[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)				
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	
WBC x10 <sup>3</sup> /µl	1	4.4	4.1	8.5	9.8	0.0460	-0.91	0.3	0.1	0.0060	1.93	
RBC x10 <sup>6</sup> /µl	1	3.31	3.3	6.61	6.4	0.0060	1.18	0.01	0.03	0.0010	-0.54	
Hb g/dl	1	10.4	10.4	20.8	23.4	0.0210	-4.38	0	0.1	0.0070	-1.28	
HCT%	1	38.7	38.2	76.9	72.05	0.1860	0.79	0.5	0.3	0.0200	0.54	
MCV-fl	1	116.9	115.6	232.5	223.5	0.4950	0.51	1.3	0.3	0.0230	2.25	
MCH-Pg	1	31.6	31.4	63	73.2	0.0760	-4.91	0.2	0.3	0.0200	-0.34	
MCHC-g/dl	1	27.4	26.8	54.2	65.3	0.1640	-1.91	0.6	0.3	0.0200	1.01	
Plt. x10 <sup>3</sup> /µl	1	363	198	561	512	1.38	1.28	165	6	0.32	27.68	
Retic %	2											

### P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs=6-8 , Poly=14-15 L=4-5, E=0-2, Mono/Promono=0-2 , B1=10-12 P.M.=22-24, Mye=18-20, Meta=20-22, Other=			
RBC Morphology	3	NORMOCYTIC NORMOCHROMIC, MICROCYTES, MACROCYTES			Poly: 25-50, Lymph: 2-7, nRBC/Mono/Eo/Blast/Pro: 0-5, Myelo: 20-35, Meta: 15-25, Baso: 0-3
Diagnosis	2	APBM			Chronic Myeloid Leukemia (Chronic Phase) - CMF CR

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist.	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10 <sup>3</sup> /μl	1	450	363	88.43	86.23	4.68	5.79	6.61	7.71
RBC x10 <sup>6</sup> /μl	1	450	363	88.43	88.98	4.96	6.06	6.34	4.68
Hb g/dl	1	450	363	85.12	90.91	6.34	4.41	8.54	4.68
HCT%	1	450	363	94.77	92.29	3.03	2.75	1.93	4.68
MCV-fl	1	450	363	97.8	92.29	0.55	3.03	1.1	4.13
MCH-Pg	1	450	363	88.98	90.91	5.51	5.23	4.96	3.03
MCHC-g/dl	1	450	363	96.97	87.05	1.38	6.06	1.38	6.06
Plt. x10 <sup>3</sup> /μl	1	450	363	90.08	90.63	5.23	5.51	4.41	3.58
ReticCount%	2	450	322	93.17	83.85	4.35	1.55	2.8	16.15
PS Assessment	3	450	352	Acceptable:75.4%,Warning Signal:24.6%,Unacceptable :0%					

**\*Comments:**

1). Among Lab (EQA) : CBC result for HB & MCH unacceptable, please check calibration/human error. Remaining results acceptable., Retic Results not reported, PS partially correct

2). Within Lab (IQA) : Difference in the CBC measurement values for PLT unacceptable, may be due to random/human error.

**Note-1:** EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----Fnd Of Report-----