



Bio Matrix Technologies Pvt. Ltd.



An ISO 9001 : 2015 Certified Company

G-114, 1st Floor, Madhura Nagar, Hyderabad - 500 038.

Ph : (040) 64525295, Mobile : +91 9553099998, E-mail : info@biomatrixtech.com, www.biomatrixtech.com

35502

INSTRUMENT CALIBRATION REPORT

CLIENT : PRANAAM HOSPITALS, MADINAGUDA, HYDERABAD
 INSTRUMENT : MICROPIPETTE
 MAKE : THERMO SCIENTIFIC SR NO : KWO3171
 MODEL : FINE PIPETTE RANGE : 5-50 µL
 LOCATION : LAB

TEST RESULTS:

Volume.	Reading in Ranges	
	5.0µl	5.0µl
1.	10.0	10.02
2.	20.0	20.04
3.	30.0	29.09
4.	40.0	39.07
5.	50.0	50.05
Mean	50.0µl	50.05µl
Acc+/-%	-0.03	0.04
CV%	0.50	0.07

TEST CONDITIONS : Ambient temperature 28 °C Humidity: 55%

STATUS : PASSED

TEST EQUIPMENT USED:

Equipment/Std's used	Model & Manufacturer (Used as a Source)	Traceability	Cal Validity
1. Weight Box	AWI	TRUE VALUE	23 - 08 - 2021
2. Weighing Balance	Sr No:020410	TRUE VALUE	23 - 08 - 2021

* EQUIPMENT USED IS TRACEABLE TO NATIONAL STANDARDS

Cal done date: 02-09-2020	Cal due date: 02-09-2021
Calibrated by: N. PETER	Approved by: N. VENKAT

