



# Bio Matrix Technologies Pvt. Ltd.



An ISO 9001 : 2015 Certified Company

G-114, 1st Floor, Madhura Nagar, Hyderabad - 500 038.

Ph : (040) 64525295, Mobile : +91 9553099998, E-mail : info@biomatrixtech.com, www.biomatrixtech.com

35501

## INSTRUMENT CALIBRATION REPORT

CLIENT : PRANAAM HOSPITALS, MADINAGUDA, HYDERABAD  
 INSTRUMENT : MICROPIPETTE  
 MAKE : THERMO SCIENTIFIC SR NO : KW1941  
 MODEL : FINE PIPETTE F3 RANGE : 100-1000 µL  
 LOCATION : LAB

### TEST RESULTS:

Volume.	Reading in Ranges	
	100.0 µl	100.2 µl
1.	300.0	300.02
2.	500.0	500.04
3.	600.0	600.02
4.	700.0	700.06
5.	1000.0	1000.05
Mean	1000.0µl	1000.5µl
Acc+/-%	-0.03	0.04
CV%	0.50	0.07

TEST CONDITIONS : Ambient temperature 28 °C Humidity: 55%

STATUS : PASSED

### TEST EQUIPMENT USED:

Equipment/Std's used	Model & Manufacturer	Traceability	Cal Validity
1. Weight Box	(Used as a Source) AWI	TRUE VALUE	23 - 08 - 2021
2. Weighing Balance	Sr No:020410	TRUE VALUE	23 - 08 - 2021

\* EQUIPMENT USED IS TRACEABLE TO NATIONAL STANDARDS

Cal done date: 02-09-2020	Cal due date: 02-09-2021
Calibrated by: N. PETER	Approved by: N. VENKAT

