



Bio Matrix Technologies Pvt. Ltd.



An ISO 9001 : 2015 Certified Company

G-114, 1st Floor, Madhura Nagar, Hyderabad - 500 038.

Ph : (040) 64525295, Mobile : +91 9553099998, E-mail : info@biomatrixtech.com, www.biomatrixtech.com

PERFORMANCE TEST REPORT FOR CENTRIFUGE

HOSPITAL NAME	PRANAAM HOSPITALS,MADINAGUDA,HYDERABAD
DEVICE	CENTRIFUGE
MANUFACTURER	R8C
MODEL	ZCIN
S NO	24884
LOCATION	LAB

PERFORMANCE STANDARDS USED

S.NO	ASSET NO.	TEST EQUIPMENT	MAKE	MODEL NO	SERIAL NO	CAL DATE	CAL DUE DATE
1	BMT/12/09	DIGITAL TACHOMETER	KUSAM-MECO	KM-2234 BL	982516	24/08/20	23/08/21
2	BMT/12/07	ELECTRICAL SAFETY ANALYZER	FLUKE	ESA-612	2111026	24/08/20	23/08/21

TEST RESULTS

Sr. No	UNIT	SET VALUE	MEASURED VALUE	DEVIATION
1	RPM	500	501	01
2	RPM	1500	1502	02
3	RPM	2500	2502	02
4	RPM	3500	3501	01
5	RPM	4500	4501	01



Bio Matrix Technologies Pvt. Ltd.



An ISO 9001 : 2015 Certified Company

G-114, 1st Floor, Madhura Nagar, Hyderabad - 500 038.

Ph : (040) 64525295, Mobile : +91 9553099998, E-mail : info@biomatrixtech.com, www.biomatrixtech.com

35246

ELECTRICAL SAFETY ANALYZER TEST RESULTS		
TEST	RESULTS	LIMITS
Line(Mains)Voltage	237	0.0 to 300.0V ac rms
Ground Wire Resistance	0.002 Ω	0.000 to 2.000 Ω
Equipment current	1.4A	0 to 20.0 A ac rms
Insulation resistance	2 M Ω	+/-2% +0.2 M Ω
Ground(Earth) Leakage	12 μ A(AC+DC)	10 to 199Ma
Chassis(Enclosure)Leakage	0.4 μ a	0.0 to 199.9 μ A
Lead to Ground and Lead to lead	0.4 μ A(AC+DC)	0.0 to 1999 μ A
Lead isolation	0.4 μ A	0.0 to 1999 μ A
Differential Leakage	0.2 μ A	0.0 to 1999 μ A
Direct Equipment Leakage	0.2 μ A	0.0 to 1999 μ A
Direct applied part Leakage	0.4Ma	0.0 to 1999 μ A
Alternative applied part patient leakage	0.01 Ω	0.0 to 1999 μ A
Point to point leakage, Voltage and Resistance	0.2 μ A,0.2v&OL Ω	+/-2% +0.2V

ELECTRICAL SAFETY TEST PASSED AS PER IEC STANDARD 60601.1

Cal done date: 02-09-2020	Cal due date: 02-09-2021
Calibrated by: N. PETER	Approved by: N. VENKAT

