



CALIBRATION REPORT

Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/20-21/00897

Page: 1 of 2

Customer Name:	JOTHI X-RAY ECG & LABORATORY	Customer Address:	1, SIVAN KOVIL STREET, THOOTHUKUDI-628002.
Location:	MICROBIOLOGY	Serial No:	7964
Description:	INCUBATOR	Tag/Asset No:	NA
Make:	KEMI	Date of Calibration:	20.09.2020
Model:	KIA 2	Date of Due:	19.09.2021

Traceability Details:

S.No	Standard Used	Make	Model	Serial No	Validity
1	Electrical Safety Analyzer	Fluke	ESA615	2244202	26-Jun-2021
2	Digital Thermo Meter	Metravi	305 XL	08000904	02-Jan-2021

Environmental Conditions and Standard Operating Procedure Details:

Temperature	24°C	Calibration Done At	ONSITE
Humidity (40-70% RH)	55%	Calibration Procedure No	CP-BM001-2019/ES CP-BM007-2019/DTM

REMARKS:

Performance Test Report : Attached

Electrical Safety Test Report : Attached

Electrical Safety Test Passed as Per IEC Standard 60601/62353

The above mentioned instruments has been calibrated using standard manufacturer recommended protocols, using equipments having traceability to National/International Standards. Partial reproduction of the certificates is not permitted.

Calibrated by:

BALAJIM

(Calibration Engineer)

Authorized Signatory:

C.SHANMUGARAJ
(Technical Manager)