



CALIBRATION REPORT

Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/20-21/00900

Page: 1 of 2

Customer Name:	JOTHI X-RAY ECG & LABORATORY	Customer Address:	1, SIVAN KOVIL STREET, THOOTHUKUDI-628002.
Location:	BIOCHEMISTRY	Serial No:	451911089
Description:	ELISA READER	Tag/Asset No:	NA
Make:	ALERE	Date of Calibration:	20-09-2020
Model:	AM2100	Date of Due:	19-09-2021

Traceability Details:

S.No	Standard Used	Make	Model	Serial No	Validity
1	Electrical Safety Analyzer	Fluke	ESA615	2244202	26-Jun-2021

Environmental Conditions and Standard Operating Procedure Details:

Temperature	23°C	Calibration Done At	ONSITE
Humidity (40-70% RH)	54%	Calibration Procedure No	CP-BM001-2019/ES

REMARKS:

Performance Test Report : Attached

Electrical Safety Test Report : Attached

Electrical Safety Test Passed as Per IEC Standard 60601/62353

The above mentioned instruments has been calibrated using standard manufacturer recommended protocols, using equipments having traceability to National/International Standards. Partial reproduction of the certificates is not permitted.

Calibrated by:


BALAJI. M

(Calibration Engineer)

Authorized Signatory:


C.SHANMUGARAJ
(Technical Manager)