



CALIBRATION REPORT

Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/20-21/00908

Page: 1 of 2

Customer Name:	JOTHI X-RAY ECG & LABORATORY	Customer Address:	1, SIVAN KOVIL STREET, THOOTHUKUDI-628002.
Location:	BIOCHEMISTRY	Serial No:	NA
Description:	MICRO PIPETTE	Tag/Asset No:	NA
Make:	MICROPET	Date of Calibration:	20-09-2020
Model:	NA	Date of Due:	19-09-2021
Range:	20-200µl		VARIABLE

Traceability Details:

S.No	Standard Used	Make	Model	Serial No	Validity
1.	Electronic Semi Micro Balance	Citizen	Cx265N	477904	08-Apr-2021

Environmental Conditions and Standard Operating Procedure Details:

ENVIRONMENTAL CONDITIONS AND STANDARD OPERATING PROCEDURE DETAILS			
Temperature	24°C	Calibration Done at	ONSITE
Humidity(40-70% RH)	55 %	Calibration Procedure No	--

REMARKS:

Test Report : Attached

Calibrated by:

BALAJIM

(Calibration Engineer)

Approved by:

C.SHANMUGARAJ
(Technical Manager)