



Result of Calibration: SBS/TR/20-21/01174

Page: 2 of 2

TEST REPORT				
Units	Nominal Values	Measured Values	Tolerance	Result
μL	5	4.9	0.5 %	PASS
μL	25	24.9	0.5 %	PASS
μL	50	50.1	0.5 %	PASS



CALIBRATION RESULT:

- Corrective maintenance required
- Removed from use
- Acceptable for use

Calibration Engineer Sign: *S. Lame*

Date: 12-10-2020



CALIBRATION REPORT

Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/20-21/01174

Page: 1 of 2

Customer Name:	IDHAYAM DIAGNOSTIC SERVICE	Customer Address:	26,DR.R.PROAD, K.K.NAGAR, HASTHINAPURAM, CHROMEPET,CHENNAI-600044
Location:	LAB	Serial No:	DX96043
Description:	PIPETTE	Tag/Asset No:	NA
Make:	ERBA	Date of Calibration:	12-10-2020
Model:	NA	Date of Due:	11-10-2021
Range:	5-50µL	VARIABLE	

Traceability Details:

S.No	Standard Used	Make	Model	Serial No	Validity
1.	Electronic Semi Micro Balance	Citizen	Cx265N	477904	08-Apr-2021

Environmental Conditions and Standard Operating Procedure Details:

ENVIRONMENTAL CONDITIONS AND STANDARD OPERATING PROCEDURE DETAILS

Temperature	24°C	Calibration Done at	ONSITE
Humidity(40-70% RH)	55 %	Calibration Procedure No	--

REMARKS:

Test Report : Attached

Calibrated by:

S. Lame
LAKSHMANAN MURUGAN.S
(Calibration Engineer)

Approved by:

C.SHANMUGARAJ
(Technical Manager)