

LAB MONTHLY SUMMARY

Lab Name **FOSTR HEALTH CARE & DIAGNOSTICS** Lab No **13328**
Month **November** Year **2020**
Constituent **HbA1c**
Group

Date of Result Entered : **17/11/2020**

Date of Report Published : **05/12/2020**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	HbA1c	OTHER IMMUNO ASSAYS	Any Analyser	169	7.21	12.55	0.90	6.6 %	-0.67	0.14

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Co-Ordinator
Christian Medical College, Vellore

***** End of Report *****