

HORIBA India Private Limited

246, Okhla Industrial Estate Phase-III,
New Delhi 110020, India
Tel : +91 (11) 4646 5000 / 4669 5001
Fax : +91 (11) 4646 5020
http : //www.horiba.com
CIN : U73100DL2006PTC153232

Date: 05/12/2019

CALIBRATION CERTIFICATE

This is to state that the hematology cell counter model **ABX YUMIZEN 500** bearing **Serial no: 711YOXH01164** at **FOSTR HEALTH CARE & DIAGNOSTICSS, Bangalore, Karnataka state**, was calibrated with Calibrator on **05/12/2019**.

The Calibration Coefficient is automatically generated after calibration. Subsequently controls were processed and found them in respective range.

Calibrator Used: **ABX MINOCAL**Lot No.: **CX 440**Expiry Date: **05/12/2019**

Sr. No	Tested Parameter	Remarks
1	Repeatability/ Precision	Passed
2	Start up/Blank Cycle	Passed
3	Calibration	Passed
4	Quality Control	Passed

Next calibration cycle is due on **04/12/2020**.

A.RAJESH,
SENIOR ENGINEER CUSTOMER SUPPORT
Horiba Medical.