



CALIBRATION CERTIFICATE

HOSPITAL NAME	FOSTR HEALTHCARE
ADDRESS	KADUBEESANAHALLI, MARATHAHALLI KARNATAKA.
ANALYSER	COBAS C111
SERIAL NO	13585
INSTALLATION PLACE	BIOCHEMISTRY DEPT.
DATE	JULY 28 , 2020

This is to certify that the above mentioned analyzer was calibrated by the following procedure as per manufacturer's recommendations

1. Analyser Rotor Init position check
2. Check Abs sensitivity 100%
3. Check Init Position Accuracy
4. Air/water calibration
5. Pipette Accuracy check

All the checks and calibration were done and system is being used under recommended conditions.
Next calibration is due on July 27, 2021.

For Roche Diagnostics India Pvt. Ltd

Manikandan K
Technical Service Specialist
Bangalore.