

ALPHA ENTERPRISES

No. 785, HARSHAKUTIRA, 4th Main, 3rd Cross, Patel Venkataswamy Road,
Subbaianapalya, Bangalore-560 033 Phone : 25456775 Mob: 9448240270
Email : alpha_ent23@yahoo.in GSTIN : 29ADWPV0639H1ZR

CALIBRATION CERTIFICATE

CERTIFICATE: No. 70-01/20
DATE: 31/08/2020

CUSTOMER: - M/S FOSTR HEALTH CARE & DIAGNOSTICS
No. 12, 2nd Floor, Panathur Main Road,
Marathahalli, Bengaluru- 560087

DESCRIPTION OF ITEM: - REFRIGERATOR (LAB)

MODEL NO: - SAMSUNG

SL.NO:- 041R4PBK2000

RANGE: 2 TO 8°C

DETAILS OF EQUIPMENT FOR CALIBRATION

Sl.No.	Description	Certified By	Certificate No.	Validity
1.	DIGITAL THERMOMETER	EDTC Govt. of India Bangalore	CR/NCAL/50993/2	11/03/2021


OBSERVATIONS:

Sl.No.	INSTRUMENT TEMP DISPLAYED(°C)	THERMOMETER DISPLAYED	TEMPERATURE DEVIATION (°C)	REMARKS
1	8.0	7.8	0.2	OK
2	6.0	5.7	0.3	OK
3	5.0	4.8	0.2	OK
4	3.0	2.9	0.1	OK

NEXT CALIBRATION: 30/08/2021

REMARKS:

1. CALIBRATION TRACEABLE TO NATIONAL STANDARDS
2. OVERALL DEVIATIONS WITHIN SPECIFIED ACCURACY +/- 5%
3. RESULTS ARE RELATED ONLY TO THE ITEM CALIBRATED



CALIBRATED BY:
Vishwas S
SERVICE ENGINEER

ALPHA ENTERPRISES

No. 785, HARSHAKUTIRA, 4th Main, 3rd Cross, Patel Venkataswamy Road,
Subbaianapalya, Bangalore-560 033 Phone : 25456775 Mob: 9448240270
Email : alpha_ent23@yahoo.in GSTIN : 29ADWPV0639H1ZR

CALIBRATION CERTIFICATE

CERTIFICATE: No. 70-01/20

DATE: 31/08/2020

CUSTOMER: - M/S FOSTR HEALTH CARE & DIAGNOSTICS
No. 12, 2nd Floor, Panathur Main Road,
Marathahalli, Bengaluru- 560087

DESCRIPTION OF ITEM: - FREEZER (Lab)

MODEL NO: - SAMSUNG

SL.NO: 041R4PBK2000

RANGE: UPTO -20°C

DETAILS OF EQUIPMENT FOR CALIBRATION

Sl.No.	Description	Certified By	Certificate No.	Validity
1.	DIGITAL THERMOMETER	EDTC Govt. of India Bangalore	CR/NCAL/50993/2	11/03/2021

OBSERVATIONS:

Sl.No.	INSTRUMENT TEMP DISPLAYED(°C)	THERMOMETER DISPLAYED	TEMPERATURE DEVIATION (°C)	REMARKS
1	-14.0	-14.3	0.3	OK
2	-16.0	-16.2	0.2	OK
3	-17.0	-17.3	0.3	OK
4	-20.0	-20.2	0.2	OK

NEXT CALIBRATION: 30/08/2021

REMARKS:

1. CALIBRATION TRACEABLE TO NATIONAL STANDARDS
2. OVERALL DEVIATIONS WITHIN SPECIFIED ACCURACY +/- 5%
3. RESULTS ARE RELATED ONLY TO THE ITEM CALIBRATED



CALIBRATED BY:

Vishwas . S
SERVICE ENGINEER

ELECTRONICS TEST & DEVELOPMENT CENTRE

(STQC Directorate, Ministry of Electronics & of Information Technology, Government of India)
 100 Feet Road, Peenya Industrial Estate, Bangalore - 560 058.
 (Phone: (080) 2839 4252/4647/4766/5992; Telefax: +91-080-23722314)
 Email: etdcbg@stqc.gov.in



Report No.: -CR/NCAL/50993-2

Page 1 of 2

CALIBRATION REPORT

Service Request No.:	50993	Date : 19-02-2020	Job No : 2
Calibrated for :	M/s, Alpha Enterprises, Banasavadi, Bangalore - 33.		
1.0 Description/Unambiguous identification of item :			
Nomenclature	Digital Thermometer with Sensor		
Make / Model	Remi / PT-100		
Serial No.	950620	Dates of Calibration:	03-03-2020 to 11-03-2020
Equipment Received on	19-02-2020	Recommended Calibration Due	11/3/2021
Condition of Item on receipt	Used	Calibrated at	In-house / Onsite
Calibration Procedure Ref.:	TM/NCAL/3	Applicable Spec.: <i>Standard/Manufacturer/User</i>	
Temperature Scale : ITS 90			
Laboratory Environmental Conditions:	Temperature: 25 ± 2 Deg. C & Humidity: 45% to 75 % RH		
Date of Issue of Calibration Report	20/3/2020		
2.0 Standards used :			
Nomenclature	Make & Model	Cal. At	Cal. Validity
Ultra cool Field Metrology Well	Fluke / 9190A	ETDC (Bg)	Source
Black Stack(SPRT Module)	Hart Scientific/2560	ETDC (Bg)	Aug-20
SSPRT	Hart Scientific/5628	ERTL(West)	Jun-20
Measurements are traceable to SI units through National Standards			

- This Calibration report refers only to the item calibrated and shall not be reproduced except in full, without the written approval from Director, ETDC, Bengaluru.
- Refer to information contained on the cover.

