



HORIBA India Private Limited

246, Okhla Industrial Estate Phase-III,
New Delhi 110020, India
Tel : +91 (11) 4646 5000 / 4669 5001
Fax : +91 (11) 4646 5020
http : //www.horiba.com
CIN : U73100DL2006PTC153232

HIN/MED/2020-2021/100701
24th Nov 2020

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer **MICROS 60** bearing serial number:
308OS93273 installed at **Mayom Hospital, Gurugram** as calibrated on 24th Nov 2020.

Calibrator : ABX MINOCAL
Lot No. : CX452
Expiry Date : 05th Dec 2020

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on **23rd Nov 2021**.



Shrish Dixit

(Head- Products & Marketing)
For **Horiba India Pvt. Ltd.**

CALIBRATION

DATE: 28/10/2020
OPERATOR: OP.1
LOT #: CX452

TIME: 15:26

RUN	WBC	RBC	HGB	HCT	PLT	MPV
1 P	9.1	4.56	13.7	39.4	237	7.3
2	9.2	4.65	14.2	40.3	233	7.6
3	9.0	4.59	13.8	39.7	235	7.6
4	9.1	4.60	14.1	39.7	250	7.5
5	9.3	4.64	14.1	40.1	239	7.7

MEAN	9.1	4.62	14.1	39.9	239	7.6
TARGET	9.2	4.54	13.9	39.9	237	7.4
C V	1.2	0.5	1.0	0.6	2.7	0.7
% CHG	0.7	1.8	1.1	0.1	0.9	2.5
OLD CAL.	1.10	0.83	1.32	1.05	1.10	0.87
CURRENT	1.11	0.81	1.31	1.05	1.09	0.85
STATUS	OK	OK	OK	OK	OK	OK