

VIEW LAB MONTHLY SUMMARY

Lab Name	RAINBOW DIAGNOSTICS (UNIT OF VIRAJ DIAGNOSTICS P LTD)	Details About Robust Analysis
Lab No	6613	Detail About Monthly Summary
Month	November	Detail about SDI
Year	2020	
Constituent Group	Chemistry I	

[Click on the analyte to view Graphical Data](#)

[All Method Result](#)

[Print](#)

Date of Result Entered : 19/11/2020

Date of Report Published : 05/12/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Any Analyser	705	387.72	9.50	36.84	382 mg/dl	-0.16	2.77
2	UREA	UREASE UV / GLDH	Any Analyser	536	57.29	10.72	6.14	88 mg/dl	5	0.53
3	CREATININE	Enzymatic Colorimetric	Any Analyser	274	8.11	10.78	0.87	8.22 mg/dl	0.13	0.11
4	T.BILIRUBIN	DIAZONIUM SALT (Colorimetric) /JENDRASSIK	Any Analyser	455	1.64	17.08	0.28	1.67 mg/dl	0.11	0.03
5	T-PROTEIN	BIURET - colorimetric	Any Analyser	572	5.13	8.51	0.44	4.9 g/dl	-0.53	0.04
6	ALBUMIN	BCG - colorimetric	Any Analyser	553	3.15	7.86	0.25	3.31 g/dl	0.65	0.02
7	CALCIUM	ARSENazo III	Any Analyser	450	10.89	8.63	0.94	11.1 mg/dl	0.22	0.09
8	PHOSPHORUS	Molybdate UV/ Phosphomolybdate complex	Any Analyser	240	3.13	13.85	0.43	3.3 mg/dl	0.39	0.06
9	CHOLESTEROL	CHOD-PAP	Any Analyser	613	111.05	8.38	9.31	109 mg/dl	-0.22	0.75
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser	566	162.70	11.55	18.80	180 mg/dl	0.92	1.58
11	HDL CHO	DIRECT METHOD / Enzymatic colorimetric	Any Analyser	415	26.88	16.71	4.49	20 mg/dl	-1.53	0.44
12	SODIUM	ISE - Indirect	Any Analyser	960	132.30	3.40	4.49	134.2 mmol/L	0.42	0.29
13	POTASSIUM	ISE - Indirect	Any Analyser	959	2.61	8.84	0.23	3.43 mmol/L	3.55	0.01
14	AST	OTHERS (Any Other Principles / Methods)	Any Analyser	225	109.45	13.01	14.24	110 U/L	0.04	1.90
15	ALT	OTHERS (Any Other Principles / Methods)	Any Analyser	228	142.95	18.96	27.11	153 U/L	0.37	3.59
	SDI Range		Interpretation							
16	ALT	PNP AMR KINETIC	Any Analyser	350	93.25	17.99	16.78	82 U/L	-0.67	1.79
	Between ± 1.0 to ± 2.0	Good.								
	Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.								
	Beyond ± 3.0	Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out
