

## LAB MONTHLY SUMMARY

Lab Name **JOTHI X-RAY AND LABORATORY**Lab No **9661**Month **November**Year **2020**Constituent Group **Chemistry II**

Date of Result Entered : 10/11/2020

Date of Report Published : 05/12/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Biosystems	224	387.58	8.09	31.37	373 mg/dl	-0.46	4.19
2	UREA I	UREASE UV / GLDH	Biosystems	134	57.13	9.88	5.64	55 mg/dl	-0.38	0.97
3	T.BILIRUBIN I	EVELYN-MALLOY	Any Analyser	79	1.64	19.10	0.31	3.6 mg/dl	6.32	0.07
4	T-PROTEIN I	BIURET - colorimetric	Biosystems	180	5.17	9.30	0.48	5.1 g/dl	-0.15	0.07
5	ALBUMIN I	BCG - colorimetric	Biosystems	180	3.21	9.16	0.29	2.9 g/dl	-1.05	0.04
6	CHOLESTEROL I	CHOD-PAP	Biosystems	210	114.08	10.52	12.00	118 mg/dl	0.33	1.66
7	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Biosystems	195	162.49	11.76	19.10	158 mg/dl	-0.24	2.74

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellpre.ac.in

Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*