



CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/THTY/CEN/D02	Date of Issue	: 11/12/2020
		Calibration Done On	: 11/12/2020
		Calibration Due On	: 10/12/2021

Customer Name	TALUK HOSPITAL, THAMARASSERY.
---------------	--------------------------------------

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Centrifuge	Remi / R – 8C Plus	ZGHN 16030	80113240	ICTC Lab

DETAILS OF MASTER EQUIPMENT USED

SI No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Tachometer	Equinox / EQ-801B	S916577	ACS/20-21/I02671-01	24/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308-CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	27°C	Relative Humidity	55 %
---------------------------------	-------------	------	-------------------	------

PERFORMANCE DATA

Set Reading	Displayed Reading	Measured Reading	Tolerance
1000 rpm	1000 rpm	1001 rpm	±5 rpm
2000 rpm	2000 rpm	2001 rpm	±5 rpm
3000 rpm	3000 rpm	3001 rpm	±5 rpm
4000 rpm	4000 rpm	4002 rpm	±5 rpm

Calibration Status	Pass	✓	Fail	
---------------------------	------	---	------	--

Calibrated By

Approved By

Prince CP (Service Engineer)

Binoj Thomas (Quality Manager)



CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/THTY/REF/D01	Date of Issue : 11/12/2020
		Calibration Done On : 11/12/2020
		Calibration Due On : 10/12/2021

Customer Name	TALUK HOSPITAL, THAMARASSERY.
---------------	--------------------------------------

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Refrigerator	Samsung / RR19R10C2SE	046RAPAM6016 47L	80113239	ICTC Lab

DETAILS OF MASTER EQUIPMENT USED

Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Thermometer	Metravi / DTM 900	10055593	ACS/20-21/I02669-01	23/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	27 °C	Relative Humidity	55 %
---------------------------------	-------------	-------	-------------------	------

PERFORMANCE DATA

Display	Measured	Tolerance
1.3 °C	1.2 °C	± 0.3 °C
2.4 °C	2.4 °C	± 0.3 °C
4.2 °C	4.3 °C	± 0.3 °C
7.2 °C	7.2 °C	± 0.3 °C

Calibration Status	Pass	✓	Fail	
---------------------------	------	---	------	--

Calibrated By

Prince CP (Service Engineer)



Approved By

Binoj Thomas (Quality Manager)



CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/THTY/VDRL/D01	Date of Issue	: 11/12/2020
		Calibration Done On	: 11/12/2020
		Calibration Due On	: 10/12/2021

Customer Name	TALUK HOSPITAL, THAMARASSERY.
---------------	--------------------------------------

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
VDRL Rotator	Rotek Plus	2030	80113238	ICTC Lab

DETAILS OF MASTER EQUIPMENT USED

Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Tachometer	Equinox / EQ-801B	10055593	ACS/20-21/I02671-01	24/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	27°C	Relative Humidity	55 %
---------------------------------	-------------	------	-------------------	------

PERFORMANCE DATA

Result	Obtained RPM
	219 rpm

Calibration Status	Pass	✓	Fail
---------------------------	------	---	------

Calibrated By

Prince CP (Service Engineer)

Approved By



Binoj Thomas (Quality Manager)



CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/THTY/CEN/D01	Date of Issue : 11/12/2020
		Calibration Done On : 11/12/2020
		Calibration Due On : 10/12/2021

Customer Name	TALUK HOSPITAL , THAMARASSERY.
---------------	---------------------------------------

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Centrifuge	Rotek	09382	67111444	ICTC Lab

DETAILS OF MASTER EQUIPMENT USED

Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Tachometer	Equinox / EQ-801B	S916577	ACS/20-21/102671-01	24/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308-CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	27°C	Relative Humidity	55 %
---------------------------------	-------------	------	-------------------	------

PERFORMANCE DATA


Set Reading	Measured Reading
1	874 rpm
2	1602 rpm
3	2223 rpm
4	2834 rpm
5	3337 rpm

Calibration Status	Pass	✓	Fail	
---------------------------	------	---	------	--

Calibrated By

Prince CP (Service Engineer)

Approved By



Binoj Thomas (Quality Manager)