



RK Technologies

Calibration and Validation Services

NABL ACCREDITED LABORATORY
AS PER STANDARD 17025 : 2017

Add : F-26, 2nd Floor, New Ashoka Vihar, XLO Point,
Ambad MIDC, Nashik - 422010.
Mobile : +91 9028646172, +91 9028777244
E-mail : rktechnologies99@gmail.com
Website : www.rktechcalibration.com

Calibration of Electro technical,
Thermal, Pressure, Mechanical,
Force & Hardness parameters.

Validation of All types of
Temperature & Humidity parameters.

CALIBRATION CERTIFICATE

Calibration Item	Micro Pipette	Certificate No	2020-12-98-20422
------------------	---------------	----------------	------------------

Date of Reciept	Date of Calibration	Next Recommended Due Date	Certificate Issue Date	Page No
11 December 2020	11 December 2020	10 December 2021	13 December 2020	01 of 01

I. Customer Name	M/S. Nidan Spciality Diagnostics Chintamani H.No:2-1-6/761 Jubli Park Signal Bhadkal Gate Aurangabad-431001
------------------	--

II. Description of Item Under Calibration :			
Instrument ID No	NSD/MP-01	Range	100 to 1000µl
Make / Model	--	Sr. No	OE505368
Type	--	Location	Lab

III Environment Condition :	
Temperature	20 to 35 °C
Relative Humidity	45 % to 75 % RH
Condition of item	Good
Location Of Calibration	In LAB

IV. Detail of Refrence Stanadard used for calibration (Traceable To National / International Standard)					
Instrument Name	ID No	Traceability (Cert No)	Date of Calibration	Valid upto	Traceability
Weight Set	RK-STD-23	C2060609/N/M&V/01	09 June 2020	8 June 2021	NABL CC 2052

V: Calibration Result :			
CalibrationPoint	Set UUC Reading	Measure Stadnard Value	Error
µl	µl	µl	µl
50	50.00	50.00	0.00
100	100.00	100.01	-0.01
500	500.00	500.03	-0.03
800	800.00	800.05	-0.05
1000	1000.00	1000.08	-0.08

VI : Note:

- 1) UUC stands for Unit Under Calibration.
- 2) Next calibration date (1 Year) mentioned in the certificate is given as per customer request
- 3) This certificate refers only to the particular item submitted for calibration
- 4) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "RK Technologies, Nashik".
- 5) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.

Calibrated By
Mr Sharad Jagdhane
Calibration Engineer



Review & Approved By
Mr Rahul Kasture
Technical Manager

***** End of Report *****

Format No. RKF - 31



RK Technologies

Calibration and Validation Services

NABL ACCREDITED LABORATORY
AS PER STANDARD 17025 : 2017

Add : F-26, 2nd Floor, New Ashoka Vihar, XLO Point,
Ambad MIDC, Nashik - 422010.
Mobile : +91 9028646172, +91 9028777244
E-mail : rktechnologies99@gmail.com
Website : www.rktechcalibration.com

Calibration of Electro technical,
Thermal, Pressure, Mechanical,
Force & Hardness parameters.

Validation of All types of
Temperature & Humidity parameters.

CALIBRATION CERTIFICATE

Calibration Item	Micro Pipette	Certificate No	2020-12-98-20423
------------------	---------------	----------------	------------------

Date of Receipt	Date of Calibration	Next Recommended Due Date	Certificate Issue Date	Page No
11 December 2020	11 December 2020	10 December 2021	13 December 2020	01 of 01

I. Customer Name	M/S. Nidan Speciality Diagnostics Chintamani H.No:2-1-6/761 Jubli Park Signal Bhadkal Gate Aurangabad-431001
------------------	---

II. Description of Item Under Calibration :			
Instrument ID No	NSD/MP-02	Range	2 to 20µl
Make / Model	--	Sr. No	OE505408
Type	--	Location	Lab

III Environment Condition :	
Temperature	20 to 35 °C
Relative Humidity	45 % to 75 % RH
Condition of item	Good
Location Of Calibration	In LAB

IV. Detail of Refrence Stanadard used for calibration (Traceable To National / International Standard)					
Instrument Name	ID No	Traceability (Cert No)	Date of Calibration	Valid upto	Traceability
Weight Set	RK-STD-23	C2060609/N/M&V/01	09 June 2020	8 June 2021	NABL CC 2052

V: Calibration Result :			
CalibrationPoint	Set UUC Reading	Measure Stadnard Value	Error
µl	µl	µl	µl
1	1.00	1.00	0.00
5	5.00	5.00	0.00
10	10.00	10.01	-0.01
15	15.00	15.02	-0.02
20	20.00	20.03	-0.03

VI : Note:

- 1) UUC stands for Unit Under Calibration.
- 2) Next calibration date (1 Year) mentioned in the certificate is given as per customer request
- 3) This certificate refers only to the particular item submitted for calibration
- 4) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "RK Technologies, Nashik".
- 5) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.

Calibrated By
Mr Sharad Jagdhane
Calibration Engineer


Review & Approved By
Mr Rahul Kasture
Technical Manager

***** End of Report *****

Format No. RKF - 31