



## CALIBRATION REPORT

### Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/20-21/01179

Page: 1 of 2

<b>Customer Name:</b>	IDHAYAM DIAGNOSTIC SERVICE	<b>Customer Address:</b>	26,DR.R.PROAD, K.K.NAGAR, HASTHINAPURAM, CHROMEPET,CHENNAI-600044
<b>Location:</b>	LAB	<b>Serial No:</b>	17583466
<b>Description:</b>	PIPETTE	<b>Tag/Asset No:</b>	NA
<b>Make:</b>	BIOHIT	<b>Date of Calibration:</b>	12-10-2020
<b>Model:</b>	NA	<b>Date of Due:</b>	11-10-2021
<b>Range:</b>	1000µl		FIXED

### Traceability Details:

S.No	Standard Used	Make	Model	Serial No	Validity
1.	Electronic Semi Micro Balance	Citizen	Cx265N	477904	08-Apr-2021


### Environmental Conditions and Standard Operating Procedure Details:

#### ENVIRONMENTAL CONDITIONS AND STANDARD OPERATING PROCEDURE DETAILS

<b>Temperature</b>	24°C	<b>Calibration Done at</b>	ONSITE
<b>Humidity(40-70% RH)</b>	55 %	<b>Calibration Procedure No</b>	--

### REMARKS:

Test Report : Attached

Calibrated by:   
LAKSHMANAN MURUGAN.S  
(Calibration Engineer)

Approved by:

C.SHANMUGARAJ  
(Technical Manager)





Result of Calibration: SBS/TR/20-21/01179

Page: 2 of 2

TEST REPORT			
Units	Nominal Values	Measured Values	Result
µl	1000	1000.1	PASS



**CALIBRATION RESULT:**

- Corrective maintenance required
- Removed from use
- Acceptable for use

Calibration Engineer Sign: *S. Lame*

Date: 12-10-2020