



## CALIBRATION REPORT

Details of the Unit Under Calibration:

Page: 1 of 2

Calibration Certificate No: SBS/CL/20-21/01178

|                |                               |                      |   |
|----------------|-------------------------------|----------------------|---|
| Customer Name: | IDHAYAM<br>DIAGNOSTIC SERVICE | Customer Address:    | 26,DR.R.PROAD, K.K.NAGAR,<br>HASTHINAPURAM,<br>CHROMEPET,CHENNAI-600044 |
| Location:      | LAB                           | Serial No:           | 14518213  |
| Description:   | PIPETTE                       | Tag/Asset No:        | NA  |
| Make:          | BIOHIT                        | Date of Calibration: | 12-10-2020  |
| Model:         | NA                            | Date of Due:         | 11-10-2021  |
| Range:         | 100-1000µl                    |                      | VARIABLE  |

Traceability Details:

| S.No | Standard Used                 | Make    | Model  | Serial No | Validity    |
|------|-------------------------------|---------|--------|-----------|-------------|
| 1.   | Electronic Semi Micro Balance | Citizen | Cx265N | 477904    | 08-Apr-2021 |

Environmental Conditions and Standard Operating Procedure Details:

| ENVIRONMENTAL CONDITIONS AND STANDARD OPERATING PROCEDURE DETAILS |      |                          |        |
|---|------|--------------------------|--------|
| Temperature   | 24°C | Calibration Done at      | ONSITE |
| Humidity(40-70% RH)   | 55 % | Calibration Procedure No | --     |

REMARKS:

Test Report : Attached

Calibrated by:

*S. Lame*

LAKSHMANAN MURUGAN.S

(Calibration Engineer)

Approved by:

C.SHANMUGARAJ  
(Technical Manager)





## SUNSHINE BIOMEDICAL SOLUTIONS

No.68, First floor, Poornagal Main Road,  
Ekkattuthangal, Chennai - 600 032, 044 - 2225 2087.

95512 66584/73387 32288

calibration@sunshinebiomedical.com



Result of Calibration: SBS/TR/20-21/01178

Page: 2 of 2

| TEST REPORT |                |                 |           |        |
|-------------|----------------|-----------------|-----------|--------|
| Units       | Nominal Values | Measured Values | Tolerance | Result |
| µl          | 100            | 100.0           | ±0.5 %    | PASS   |
| µl          | 500            | 500.1           | ±0.5 %    | PASS   |
| µl          | 1000           | 1000.1          | ±0.5 %    | PASS   |



### CALIBRATION RESULT:

- Corrective maintenance required
- Removed from use
- Acceptable for use

Calibration Engineer Sign: *S. Lamey*

Date: 12-10-2020

SERVICE - CHENNAI.