



**CALIBRATION REPORT**

**Details of the Unit Under Calibration:**

**Calibration Certificate No: SBS/CL/20-21/01180**

Page: 1 of 2

<b>Customer Name:</b>	IDHAYAM DIAGNOSTIC SERVICE	<b>Customer Address:</b>	26,DR.R.PROAD, K.K.NAGAR, HASTHINAPURAM, CHROME PET,CHENNAI-600044
<b>Location:</b>	LAB	<b>Serial No:</b>	13582032
<b>Description:</b>	PIPETTE	<b>Tag/Asset No:</b>	NA
<b>Make:</b>	BIOHIT	<b>Date of Calibration:</b>	12-10-2020
<b>Model:</b>	NA	<b>Date of Due:</b>	11-10-2021
<b>Range:</b>	10-100µL	VARIABLE	

**Traceability Details:**

S.No	Standard Used	Make	Model	Serial No	Validity
1.	Electronic Semi Micro Balance	Citizen	Cx265N	477904	08-Apr-2021

**Environmental Conditions and Standard Operating Procedure Details:**

*ENVIRONMENTAL CONDITIONS AND STANDARD OPERATING PROCEDURE DETAILS*

<b>Temperature</b>	24°C	<b>Calibration Done at</b>	ONSITE
<b>Humidity(40-70% RH)</b>	55 %	<b>Calibration Procedure No</b>	--

**REMARKS:**

Test Report : Attached

Calibrated by:

*S. Lamey*

LAKSHMANAN MURUGAN.S

(Calibration Engineer)

Approved by:

C.SHANMUGARAJ

(Technical Manager)





Result of Calibration: SBS/TR/20-21/01180

Page: 2 of 2

TEST REPORT				
Units	Nominal Values	Measured Values	Tolerance	Result
$\mu\text{L}$	10	9.9	0.5 %	PASS
$\mu\text{L}$	50	49.9	0.5 %	PASS
$\mu\text{L}$	100	100.1	0.5 %	PASS



**CALIBRATION RESULT:**

- Corrective maintenance required
- Removed from use
- Acceptable for use

Calibration Engineer Sign: *S. Iame*

Date: 12-10-2020