



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



Certificate No.
CC-2239

CALIBRATION CERTIFICATE

Customer Name And Address	M/S : Rainbow Diagnostics	Certificate No.	CC2239206000042601F
	Unit of	Service request No & date	A/01/11/12/2020
	SC-45, C Block, Near Union Bank,	Date of calibration	11/12/2020
	Shalimar Garden Extn. II, Sahibabad,	Valid up to	10/06/2021
	Ghaziabad	Date of issue	11/12/2020

Instrument Details

Instrument name	Centrifuge Machine	Sr. No.	-----
Make	Remi	Location	-----
Range / Size	0 to 5000 rpm	Accuracy	-----
Least Count	As per Range	Visual Inspection	OK
I.D. No.	CDW - 06		

Detail of reference standards & Major equipments used

Equipment Name	Digital Tachometer		
Make	LT Lutron		
Model / SR No.	-----/AC92792		
Certificate No.	CC294020EL0001124F		
Calibration Validity	22/11/2021		
Calibration by	Digital Calibration Technology		

Environmental Condition	Temperature	20±2°C	Calibration Reference	IS: 12508
	Relative Humidity	50±10 %	Calibration Procedure	CP-37


Calibration Results

Serial No.	Parameter	Standard value in (rpm)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	Position - 1	1785	±1 %rdg
02.	Position - 2	2263	
03.	Position - 3	3175	
04.	Position - 4	3585	
05.	Position - 5	4826	

Remarks:

- ❖ This result of calibration refer to only to the particular item submitted for calibration.
- ❖ The above result are valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. are scientific purpose only and should not used for commercial.

Calibrated By
(Calibration Engg./TM)
(MANPAL)
Form No.-32



---End of Report---

Approved By
(QM/TM)
(NEERAJ YAGI)
Page No. 1 of 1